

Medical Coverage Policy



Effective Date..... 1/15/2021
Next Review Date..... 1/15/2022
Coverage Policy Number 0536

Prosthetic Devices

Table of Contents

Overview	1
Coverage Policy	2
General Background	5
Medicare Coverage Determinations	16
Coding/Billing Information	16
References	22

Related Coverage Resources

[Intraocular Lens Implant](#)
[Male Sexual Dysfunction Treatment: Non-Pharmacogenic Treatment of Gender Dysphoria](#)

INSTRUCTIONS FOR USE

The following Coverage Policy applies to health benefit plans administered by Cigna Companies. Certain Cigna Companies and/or lines of business only provide utilization review services to clients and do not make coverage determinations. References to standard benefit plan language and coverage determinations do not apply to those clients. Coverage Policies are intended to provide guidance in interpreting certain standard benefit plans administered by Cigna Companies. Please note, the terms of a customer's particular benefit plan document [Group Service Agreement, Evidence of Coverage, Certificate of Coverage, Summary Plan Description (SPD) or similar plan document] may differ significantly from the standard benefit plans upon which these Coverage Policies are based. For example, a customer's benefit plan document may contain a specific exclusion related to a topic addressed in a Coverage Policy. In the event of a conflict, a customer's benefit plan document always supersedes the information in the Coverage Policies. In the absence of a controlling federal or state coverage mandate, benefits are ultimately determined by the terms of the applicable benefit plan document. Coverage determinations in each specific instance require consideration of 1) the terms of the applicable benefit plan document in effect on the date of service; 2) any applicable laws/regulations; 3) any relevant collateral source materials including Coverage Policies and; 4) the specific facts of the particular situation. Coverage Policies relate exclusively to the administration of health benefit plans. Coverage Policies are not recommendations for treatment and should never be used as treatment guidelines. In certain markets, delegated vendor guidelines may be used to support medical necessity and other coverage determinations.

Overview

This Coverage Policy addresses prosthetic devices. Prosthetic devices are defined as fabricated items designed as replacements for missing body parts.

The policy statements below provide medical necessity criteria, including functional level requirements where applicable, and coding information for the following:

- [General Criteria for any Prosthetic Device](#)
- [External Facial Prosthetic Device](#)
- [Upper Limb Prosthetic Device \(Myoelectric\)](#)
- [Lower Limb Prosthetic Device \(Microprocessor-controlled, Powered-microprocessor controlled, Vacuum Suspension System\)](#)
- [Repair and Replacement](#)

For information regarding medical necessity criteria for any other prosthetic device please reference the applicable Cigna Medical Coverage Policy:

- [Breast Reconstruction Following Mastectomy or Lumpectomy](#)
- [Intraocular Lens Implant](#)
- [Male Sexual Dysfunction Treatment: Non-Pharmacogenic](#)

- [Treatment of Gender Dysphoria](#)

Coverage Policy

Coverage for prosthetic devices varies across plans. Please refer to the customer's benefit plan document to determine benefit availability and the terms and conditions of coverage.

Microprocessor-controlled/computer-controlled/myoelectric devices are considered a type of power enhancement/controlled device.

GENERAL CRITERIA FOR A PROSTHETIC DEVICE

Functional Levels

Medical necessity for a lower limb prosthetic appliance is based on an individual's functional ability when using the prosthetic device. Functional ability is based on the following classification levels:

- **Level 0:** Does not have the ability or potential to ambulate or transfer safely with or without assistance and prosthesis does not enhance his/her quality of life or mobility.
- **Level 1:** Has the ability or potential to use prosthesis for transfers or ambulating on level surfaces at fixed cadence; typical of the limited and unlimited household ambulator.
- **Level 2:** Has the ability or potential for ambulating with the ability to traverse environmental barriers such as curbs, stairs or uneven surfaces; typical of the limited community ambulator.
- **Level 3:** Has the ability or potential for ambulating with variable cadence; typical of the community ambulator who has the ability to traverse most environmental barriers and may have vocational, therapeutic, or exercise activity that demands prosthetic utilization beyond simple locomotion.
- **Level 4:** Has the ability or potential for prosthetic ambulating that exceeds basic ambulating skills, exhibiting high impact, stress, or energy levels; typical of the prosthetic demands of the child, active adult, or athlete.

The following prosthetic devices are considered medically necessary when used to replace a missing or nonfunctional body part and when applicable medical necessity criteria listed below is met (Please note: prior authorization requirements may apply):

- External facial (e.g., nose, ear, midfacial, orbital, upper facial, hemifacial)
- Eye prosthesis (e.g., internal ocular, scleral shell)
- Lower extremity (e.g., foot, ankle, above/below knee)
- Upper extremity (e.g., finger, hand, wrist, above/below elbow, shoulder)
- Terminal devices, such as hands or hooks

Accessories to a prosthetic device are considered medically necessary when the accessory is required for the effective use of the prosthesis.

Not Medically Necessary

The following prosthetic devices are each considered not medically necessary:

- a lower limb prosthetic device for functional level 0
- additions/components that are not required for the effective use of the device
- consumable supplies for the care of prosthetic device (e.g., cosmetics, creams, cleansers, skin barrier wipes)
- prosthetic devices or additions/components not required for participation in normal activities of daily living, including those that are chiefly for convenience, for participation in recreational activities, or that

otherwise exceed the medical needs of the individual (e.g., back-up/duplicate prosthetic devices, waterproof leg prosthesis [e.g., Water Leg, used for showering, swimming])

IRIS PROSTHESIS

An iris prosthesis (HCPCS code C1839) for the treatment of full or partial aniridia is considered experimental, investigational or unproven.



EXTERNAL FACIAL PROSTHESIS

An external facial prosthesis (HCPCS code L8040, L8041, L8042, L8043, L8044, L8045, L8046, L8047 and L8048) is considered medically necessary when the prosthesis is prescribed to compensate for the loss or absence of facial tissue as a result of disease, injury, surgery or congenital defect.

A duplicate external facial prosthesis is considered a convenience item and is considered not medically necessary.

Each of the following supplies related to the care of, and/or application or removal of, an external facial prosthesis is a consumable item specifically excluded under most benefit plans and considered not medically necessary:

- cosmetics
- skin creams
- skin cleansers
- adhesives
- adhesive remover
- skin barrier wipes
- tape

UPPER LIMB: MYOELECTRIC PROSTHETIC DEVICE

If a benefit is available for an upper limb myoelectric device the following medical necessity criteria apply.

An upper limb myoelectric prosthetic device is considered medically necessary for an individual with an amputation or congenital absence of a portion of an arm (e.g., hand, forearm, elbow) when ALL of the following criteria are met:

- The individual has sufficient cognitive ability to successfully utilize a myoelectric prosthetic device.
- The remaining musculature of the arm(s) contains the minimum microvolt threshold to allow operation of a myoelectric prosthetic device.
- A standard body-powered prosthetic device cannot be used or is insufficient to meet the functional needs of the individual in performing activities of daily living

An upper limb sensor and myoelectric controlled prosthetic device with simultaneous multiple degrees of freedom (e.g., LUKE [Life Under Kinetic Evolution] Arm) is considered experimental, investigational or unproven.

LOWER LIMB: MECHANICAL (NON-POWERED, NON MICROPROCESSOR)

The following lower limb additions and/or components are considered medically necessary when the individual is functional level 3 or greater and medical necessity criteria has been met for the base device:

- A flex-walk system or equal, all lower extremity prosthesis (HCPCS code L5981)

- a single axis, fluid swing and stance phase control (HCPCS L5828)
- a fluid stance extension, dampening feature, with or without adjustability (HCPCS L5848)

An adjustable stance flexion feature (HCPCS L5845) is considered medically necessary when the individual is functional level 1 or greater and medical necessity criteria has been met for the base device.

A high activity knee control frame (HCPCS code L5930) is considered medically necessary for an individual who is functional level 4 and medical necessity criteria has been met for the base device.

LOWER LIMB MICROPROCESSOR-CONTROLLED PROSTHETIC DEVICE

If a benefit is available for a microprocessor-controlled/computer-controlled lower limb prosthetic, the following medical necessity criteria apply.

Any of the following microprocessor-controlled prosthetics, including additions/components that are required for the effective use of the device (and consistent with the user's functional level), are considered medically necessary when the individual is functional level 3 or greater:

- a microprocessor-controlled ankle-foot prosthetic (HCPCS code L5973) for a transtibial amputee (below-the-knee)
- a microprocessor-controlled knee prosthetic (HCPCS code L5856, L5857, L5858) for a knee disarticulation amputee or a transfemoral amputee (above-the-knee)
- a combination microprocessor-controlled prosthetic/system (e.g., SYMBIONIC® LEG 3, LiNX®), when a microprocessor-controlled prosthetic knee alone is inadequate to meet the functional needs of the individual (e.g., continued knee/foot instability due to environmental/anatomical barriers)

A microprocessor-controlled prosthetic is considered experimental, investigational or unproven for any other indication.

An osseointegrated/osseoanchored lower limb prosthetic device is considered experimental, investigational or unproven.

LOWER LIMB: POWERED MICROPROCESSOR-CONTROLLED PROSTHETIC DEVICE

If a benefit is available for a powered or power-enhanced lower limb prosthetic, the following medical necessity criteria apply.

An endoskeletal knee-shin system (addition to a lower limb device) with powered and programmable flexion/extension assist control, including any type of motor(s) (HCPCS code L5859) (e.g., Össur Power Knee™) is considered medically necessary when ALL of the following criteria have been met:

- Individual has a swing and stance phase type microprocessor controlled (electronic) knee (HCPCS L5856)
- Is K3 functional level only*
- Has a documented comorbidity of the spine and/or sound limb affecting hip extension and/or quadriceps function that impairs K-3 level function with the use of a microprocessor-controlled knee alone

***Note:** Coverage of this device is limited to individuals who are Functional Level 3; the device is not intended for high impact activity, sports, excessive loading, or heavy duty use.

The following powered prosthetic devices are each considered experimental, investigational or unproven:

- a microprocessor-controlled ankle foot prosthetic with power assist (e.g., BiOM® Ankle, emPOWER™ Ankle [HCPCS L5973, L5969])
- a powered lower limb prosthetic for any other indication

LOWER LIMB: VACUUM SUSPENSION SYSTEM

A vacuum suspension system (e.g., vacuum-assisted socket system [VASS™]) (HCPCS code L5781, L5782) is considered medically necessary to control residual limb volume when there is contraindication to or failure of other socket-suspension systems (e.g., mechanical, passive suction) to adequately secure the limb to the prosthesis.

REPAIR AND REPLACEMENT

Repair and/or replacement of a medically necessary prosthetic device is considered medically necessary for EITHER of the following indications:

- when anatomical change or reasonable wear and tear renders the item nonfunctional and the repair will make the equipment usable.
- when anatomical change or reasonable wear and tear renders the item nonfunctional and nonrepairable.

General Background

PROSTHETIC DEVICE

A prosthesis is an artificial device used to replace a missing body part and is intended to restore normal function.

The following services and items are typically included in the allowance for a prosthetic device:

- the evaluation and fitting of the prosthesis
- the cost of base component parts and labor, as described in HCPCS base codes
- the repairs due to normal wear and tear during the 90-day period following the date of delivery
- adjustments of the prosthesis or the prosthetic component made when fitting the prosthesis or component and for 90 days from the date of delivery, when the adjustments are not necessitated by changes in the underlying tissue or the patient's functional ability

Prosthetic devices are secured or retained in place by harnesses or belts, by suction, or using anatomical structures; some devices such as facial prosthetics are held in place with the use of a skin adhesive. Additionally, devices may be held in place by implants, such as bone integrated titanium implants.

U.S. Food and Drug Administration (FDA)

Prosthetic devices are subject to regulation by the FDA as medical devices. Prosthetic accessories and limb components are classified by the FDA as Class I devices.

IRIS PROSTHESIS

An iris prosthesis is an implanted device recommended for treatment of partial or complete aniridia. Aniridia is absence of the iris and may be associated with visual conditions such as glare, photophobia, glaucoma, corneal opacification, and/or cataract formation. The degree of vision loss varies. Treatment generally consists of contact lenses with iris prints and tinted eyeglasses. The prosthetic iris device is made out of foldable medical grade silicone which is then custom-sized and colored for each individual. The iris prosthetic is implanted surgically through a small incision, it is then unfolded, the edges are smoothed out and it is then held in place by anatomical structure of the eye or using sutures. It may be placed in the ciliary sulcus without sutures when there is a pre-existing intraocular lens, implanted into the capsular bag with a new intraocular lens, or can be sutured to the sclera, with or without an IOL. The device allegedly reduces sensitivity to light while improving the appearance of the eye and visual acuity. Implant insertion can be done alone or in combination with cataract or lens fixation surgery.

The CustomFlex™ Artificial Iris (Clinical Research Consultants, Inc., Cinn., OH [HumanOptics]) received premarket approval (P170039) by the U.S. Food and Drug Administration (FDA) in May 2018 as an artificial iris

intended for use in children and adults for the treatment of full or partial aniridia resulting from congenital aniridia, acquired defects, or other conditions associated with full or partial aniridia. The device is available with or without embedded fiber mesh for implantation, and may or may not be sutured. The FDA is requiring a post approval study to evaluate long term safety outcomes up to three years postoperatively for adults and five years for pediatric subjects.

There is a growing body of evidence in the peer-reviewed scientific literature evaluating use of the artificial iris. In general, sample populations are small, studies are retrospective, study populations are heterogeneous, and surgical techniques vary precluding generalization of overall safety and efficacy. Spitzer et al (2016) published the results of a retrospective case series involving 34 subjects who received a customized silicone iris prosthesis (Artificial Iris, HumanOptics, Germany) after severe globe injury with total or sub-total iris loss. The Artificial Iris is a customized, silicone prosthetic iris made from silicone material. The median follow-up was 24 months (range 12.0-48.8). Five patients (15%) had pre-existing glaucoma and eight patients (24%) had pre-existing hypotony. Mean visual acuity prior to artificial iris implantation was 1.1 logMAR (range 0.3-2.6). At 12 months after surgery 14 subjects had VA improvement between 0.2 and 2.1 logMAR units (41%), 11 subjects had a VA change of less than 0.2 logMAR units (32%), and nine subjects had a reduction of VA between 0.2 and 1.4 logMAR units (27%). Visual acuity 12 months after surgery was 1.4 logMAR (range 0.2-2.6); median VA was unchanged. Complications included newly diagnosed glaucoma (9%) and hypotony (9%), persisting intraocular inflammation (8.8%), macular edema (11.8%), and corneal endothelial decompensation requiring corneal transplantation (18%). Patients' satisfaction increased by reducing photophobia and enhanced cosmetic appearance; 15 subjects had reduced subjective glare and while a majority of subjects were satisfied with functional and cosmetic results (80%), three continued to have persistent glaring or deteriorating vision and were not satisfied. Limitations of the study small sample population, short-term outcomes, lack of a statement regarding subjective discomfort due to glaring from 14 subjects (information was only available for 20 subjects at follow-up).

Mayer and colleagues (2016) reported results of a prospective case series investigating functional results and patient satisfaction after surgical iris reconstruction. Thirty-seven consecutive patients with traumatic iris defects, presenting from 2011 through 2014 who underwent pupillary reconstruction with a new artificial iris implant (Artificial Iris, HumanOptics, Germany), were included in the study. The main outcome measures included change of best-corrected visual acuity (BCVA), intraocular pressure (IOP), pupillary aperture, glare, contrast sensitivity, endothelial cell density, anterior chamber depth, anterior chamber angle, and patient satisfaction. Thirty-two eyes of 32 patients (mean age, 52.9±16.0 years) were included. After implantation and during follow-up, BCVA and IOP did not change significantly (BCVA, 0.77±0.62 logarithm of the minimum angle of resolution [logMAR] preoperatively vs. 0.68±0.64 logMAR 1 month postoperatively [P = 0.792]; (IOP, 14.94±3.55 mmHg preoperatively vs. 17.72±5.88 mmHg 1 month postoperatively [P = 0.197]). The pupillary aperture was reduced significantly (42.11±20.1 mm² to 8.7±0.3 mm²; P < 0.001). Contrast sensitivity increased significantly (0.80±0.51 to 0.93±0.49; P = 0.014). Endothelial cell count revealed a significant decrease postoperatively (1949±716 per 1 mm² to 1841±689 per 1 mm²; P = 0.003). Anterior chamber depth (4.03±1.06 mm preoperatively vs. 4.29±0.70 mm postoperatively; P = 0.186) and angle (43.2±13.5° preoperatively vs. 40.5±10.8° postoperatively; P = 0.772) showed no significant differences. Subjective impairment through glare (9.12±1.62 preoperatively vs. 3.07±2.29 postoperatively; P < 0.001) and cosmetic disturbance (6.33±3.21 preoperatively vs. 1.58±0.86 postoperatively; P < 0.001) improved significantly. Overall patient satisfaction was 8.91±1.51 of 10 points on an analog scale. The authors concluded that the implantation of the artificial iris is an effective therapeutic option for the treatment of traumatic iris defects and results in an "individual, aesthetically appealing, and good functional outcome in addition to high patient satisfaction". Limitations of the study as noted by the authors include five subjects excluded from follow-up, and inclusion of subjects with varying iris defects.

Rickman et al (2016) reported a retrospective interventional case series of 34 patients who received an artificial iris between 2004 and 2013 using the Artificial Iris (HumanOptics, Germany). Only eyes with a minimum follow-up period of 2 years were included, subjects ranged in age from 28-85 years. Indications for treatment were congenital, traumatic, or iatrogenic complete or partial aniridia. The artificial iris was implanted either with or without embedded fiber mesh for partial or full prostheses. Mean followup was 50.0 months (SD ±18.9 months). Repositioning of prostheses was not required in any of the 34 cases. In cases of keratopathy (17.6 %) visual function increased from baseline mean 1.6 logMAR (SD ±0.7) to 1.2 logMAR (SD ±0.7) after artificial iris implantation. The remaining iris tissue darkened during the follow-up in 23.5 % (83.3 % with and 10.7 % without mesh), 8.8 % developed glaucoma (50 % with and 0 % without mesh) and 14.7 % needed consecutive surgery

after prostheses implantation (50 % with and 7.1 % without mesh). In three out of seven trauma cases (42.9 %) silicone oil was spilled into the anterior chamber after 2.5 years, on average. When the VA at baseline was compared to the final examination, 16 eyes gained two or more VA lines, 15 eyes remained stable and 3 eyes lost two or more VA lines. There was no significant difference in the mean IOP when baseline was compared to final examination. According to the authors, the artificial iris prosthesis revealed a good clinical outcome in terms of long-term stability, cosmetic appearance and visual function. Limitations noted by the authors included a wide range of aniridia causes and variation in disease and management. Therefore, direct correlation of the success rate and the surgical technique is not firmly established. Furthermore, the authors acknowledged long-term complications such as glaucoma, over-pigmentation of the remaining iris tissue, and need for a secondary surgery are significantly associated with implants with integrated fiber mesh, however not to implants without mesh.

Mostafa and associates (2018) evaluated the limitations and benefits of the BrightOcular prosthetic artificial iris (Stellar Devices) device in management of aniridia associated with aphakia or cataract. Designed as a retrospective study, the authors evaluated 5 eyes of 4 patients (ages 12, 13, 28 and 34 years) who underwent implantation of the BrightOcular iris prosthesis (Stellar Devices) for total or partial aniridia. Similar to the HumanOptics prosthesis, this device is silicone, yet not FDA approved. The study group included 2 eyes of 1 patient with congenital aniridia associated with congenital cataract, and 3 eyes with traumatic aniridia (1 with subluxated cataractous lens and 2 with aphakia). The iris prosthesis was implanted after a 3-piece acrylic intra-ocular lens (IOL) was implanted in all cases. Measured outcomes included intra-operative and post-operative complications, and the cosmetic satisfaction and evaluation of the clinical course for at least six months. Uncorrected distance VA and best-corrected distance visual acuity (BCVA) improved for all subjects. All patients had a transient corneal edema that resolved within the 1st post-operative week. Only the patient with congenital aniridia had a permanent increase in IOP and developed a band keratopathy throughout a 2-year follow-up period. The prosthesis was well-centered in all eyes except for 1 case that needed scleral suture fixation after 3 months. One case required scleral suturing due to intraoperative displacement. In the authors opinion both cases were the result of improper sizing of the device. It was reported all subjects had a satisfactory cosmetic appearance, and improvement in glare and halos. The authors concluded that the BrightOcular iris prosthesis was a safe and useful tool to correct aniridia associated with pseudophakia or aphakia. In addition, more research is required to determine the best means of sizing the implant and to address the problem of post-operative IOP rise; further studies should also examine the safety of the prosthesis in clear phakic eyes. Limitations of the study include the small sample population and retrospective study design.

Mayer, et al. (2018) retrospectively evaluated the learning curve of the implantation surgery for the iris prosthesis and potential complications. A total of 51 subjects were implanted with the Artificial Iris, (HumanOptics, Germany), follow-up occurred at least three months post procedure and extended to a maximum of four years. Complications were grouped into categories of none, mild (with full recovery) or moderate (without full recovery) and severe (required surgical intervention). The overall complication rate was 25.5% (13/51 subjects). Mild complications included recurrent bleeding with rise in IOP (n=1), slight but stable iris deviation (n=2), capsular fibrosis (n=2); moderate complications included suture cutting through the residual iris (n=1), new onset glaucoma (n=3), and corneal decompensation (n=5); severe complications included iris suture loosening (n=2), and dislocation (n=3), synechiae (n=2), glaucoma (n=2), and corneal decompensation (n=5), with need for surgery, cystoid macular edema (n=3) and retinal detachment (n=1). The complication rate decreased from 83.3% in the first year to 13.3% in the fourth year. The author group concluded implantation of the artificial iris implant requires significant surgical experience, should be limited to specialized centers, and requires careful postoperative management to detect unexpected adverse events.

Yoeruek and Bartz-Schmidt (2019) reported the results of a small case series involving five subjects with traumatic aniridia, combined with aphakia and corneal scars or graft failure, who received an intraocular lens attached to a customized silicone iris prosthesis (Artificial Iris, HumanOptics). The mean age of the subjects was 46.2 years and the mean follow-up was 24.6 months. The mean BCVA improved from 1.36 logMAR before surgery to 0.78 logMAR after surgery during the follow-up. Data on glare and photophobia was available for three subjects; in three glare sensation was reduced. Postoperative complications included one graft failure during the first year after surgery. Three subjects had glaucoma prior to surgery; two were able to be controlled sufficiently postoperatively. There was no new cases of glaucoma postoperatively. At the last follow-up visit, the artificial iris-IOL complex was well-centered with good positioning in all cases. The authors concluded that

management of post-traumatic aniridia combined with aphakia and corneal scars or graft failure by haptic fixation of a foldable IOL on an artificial iris combined with a simultaneous keratoplasty appeared to be a promising approach, which allowed to correct a complex lesion with a less traumatic and faster procedure. The study is limited by the small sample size, retrospective design and short term follow-up.

Mayer and colleagues (2019) reported the results of a single center case series to evaluate the effect of an artificial iris implant on a remnant iris ($n = 42$). Morphologic evaluation was carried out over 24 ± 14 months. Main outcome measures included remnant pupillary aperture, iris color, VA, IOP, and endothelial cell count (ECC). Retraction syndrome, manifest by progressive enlargement of the pupil and retraction of the residual iris, was detected in seven of 42 (16.7%) eyes following implantation of the artificial iris prosthesis. Residual iris aperture dilated from $36.6 \pm 15.4 \text{ mm}^2$ pre-operatively to $61.1 \pm 12.5 \text{ mm}^2$ one year post-operatively (66.9 % increase). In 5 of 7 affected eyes, the artificial iris had been implanted into the ciliary sulcus; in 2 eyes it had been sutured to the sclera. A total of 4 of the 7 subjects presented with remarkable complications: 2 eyes needed glaucoma shunt surgeries owing to pigment dispersion; 1 suffered from recurrent bleeding; and in 1 case artificial iris explantation was performed owing to chronic inflammation and elevated intraocular pressure. Anterior chamber depth (ACD) and angle, ECC, and VA did not change in this cohort. Changes in color were not observed in the remnant iris. The authors concluded that the implantation of an artificial iris prosthesis could lead to a residual iris retraction syndrome as a late complication. It was likely that residual iris was trapped in the fissure between the artificial iris and the anterior chamber angle, preventing further pupil constriction. Another possibility noted by the authors could be the result of a constriction or atrophy of the residual iris. Due to the small sample population the authors were unable to determine statistical comparisons regarding different implantation methods. They concluded that with increased use of the artificial iris more cases of iris retraction syndrome may be detected in the future.

Figueiredo and Snyder (2020) retrospectively evaluated the safety and effectiveness of the CustomFlex device when used to treat photic symptoms in individuals with congenital aniridia ($n=50$ subjects, 96 eyes). Mean follow-up was 44 months (36 ± 36 months). Measured outcomes included pre and post-operative data regarding corrected distance visual acuity (CDVA), subjective photophobia and glare, keratopathy, glaucoma, IOP, glaucoma drops, and other comorbid pathologies. Additional postoperative data regarding postoperative complications, prosthesis decentration, and further surgeries was also collected. In all cases, additional procedures were performed at the time of implantation, including phacoemulsification, intraocular lens (IOL) implantation repositioning or replacement, limbal relaxing incision, keratectomy (superficial and lamellar) or vitrectomy. Intraoperative complications were reported in 14 eyes (14.6%). A total of 95.7% (89/93) reported a reduction in photophobia symptoms, 3.2% (3/93) reported no change in symptoms and 1.1% (1/93) reported worsening of symptoms. Similarly, subjective reporting of glare indicated a reduction of symptoms in 95.2% of subjects (79/83), 3.6% (3/83) reported no change in symptoms and 1.2% (1/83) reported worsening of symptoms. When individuals could not reliably report their symptoms, family member observations of behaviors was used to gauge functional improvement in photic symptoms. When preoperative visual acuity was compared to best achieved postoperative visual acuity, it was found that 72 eyes (75.0%) gained at least 2 lines and 24 eyes (25.0%) stayed within 2 lines, whereas no eye lost 2 or more lines. When compared with last measured visual acuity 58.3% (56) of the eyes improved 2 or more lines, 32.3% (31) of the eyes stayed within two lines of preoperative measurements, and 9.4% (9) of the eyes dropped two or more lines. The declines in the VA in the postoperative period were attributed to underlying comorbidities, which included worsening of the ocular surface, aniridia fibrosis syndrome, retinal detachment, and posterior capsule opacification. Aniridic keratopathy, which was present in 84.4% (81) of the eyes preoperatively, was present in 85.4% (82) at last visit (28.4% [23] of the eyes with preoperative keratopathy had progression of the disease). Aniridic glaucoma was present in 33.3% (32) of the eyes preoperatively in comparison with 51.0% (49) of the eyes at last visit (53.1% [17] of the eyes with preoperative glaucoma had progression of the disease). Additional complications included aniridia fibrosis syndrome (AFS) (3.1%), prosthesis decentration (9.4%), choroidal folds/effusion secondary to ocular hypotony (2.1%), retinal detachment (1.0%), cystoid macular edema (1.0%) and vitreous hemorrhage (1.0%). Overall 33.3% (32) eyes required additional surgical intervention. In the authors opinion individuals with congenital aniridia syndrome present with highly complex eyes which require an individualized approach and long-term follow-up. Limitations noted by the authors included significant heterogeneity related to aniridic pathology within the group.

The National Institute for Health and Care Excellence (NICE) published interventional procedures guidance for artificial iris insertion as treatment for acquired aniridia (NICE, 2020). NICE reviewed evidence consisting of one non-randomized comparative trial, seven case series, and one case report. The primary efficacy outcomes included reduction in symptoms of glare, improvement in visual acuity, quality of life and other patient-reported outcomes. Key safety outcomes included need for explantation, infection, worsening visual acuity, glaucoma, and implant displacement. Within this document NICE concluded the “evidence on the safety and efficacy of artificial iris implant insertion for acquired aniridia is limited in quantity and quality. Therefore, this procedure should only be used with special arrangements for clinical governance, consent, and audit or research.”

Hayes published a search and summary report in January 2019 and concluded the evidence was insufficient to evaluate safety and efficacy of the device for treatment of aniridia. Evidence reviewed by Hayes included two prospective uncontrolled studies, two retrospective uncontrolled studies, four case series, and a surgical technique review (Hayes, 2019).

Although promising, evidence in the peer reviewed scientific literature evaluating use of the artificial iris prosthesis has not firmly established safety and efficacy of the device. Professional society statements regarding use of the device as treatment for aniridia from the American Academy of Ophthalmology and American Association for Pediatric Ophthalmology were not found. Within the clinical studies several authors have reported high complications rates, both intra and post-operatively. As a result strong evidence based conclusions regarding safety and efficacy cannot be made. Additional clinical studies with longer followup are needed to evaluate use of the device and impact on health outcomes.

EXTERNAL FACIAL PROSTHESIS

External facial prostheses are used to replace lost or absent facial tissue that is the result of disease, injury, surgery or a congenital defect or they may be considered an alternative to reconstructive surgery. An external device is usually made from silicone materials and requires frequent removal and cleaning while a surgically implanted prosthetic device is typically removed and cleaned less often. The function of the external prosthesis is to protect exposed tissues, cover exposed cavities, and restore physical appearance.

Common types of external facial prostheses include the following:

- auricular (ear) - restores all or part of the ear, function includes directing sound into the auditory canal; supporting eyeglasses and acting as a hearing aide if required.
- nasal (nose) - restores all or part of the nose and may include the nasal septum; functions to direct airflow to the nasopharynx and may also provide support for eyeglasses
- midfacial (nose and adjacent tissues) - restores part or all of the nose and significant adjacent facial tissue/structures, does not include the orbit or any intraoral maxillary prosthesis; adjacent facial tissue/structures include one or more of the following: soft tissue of the cheek, upper lip, or forehead.
- orbital (orbit/eyelids) - restores the eyelids and the hard and soft tissue of the orbit, may include the eyebrow; functions to house the artificial eye, does not include the ocular prosthesis
- upper facial (orbit and adjacent tissues) - restores the orbit, plus significant adjacent facial tissue/structures, does not include the nose, any intraoral maxillary prosthesis or ocular prosthesis; adjacent facial tissue/structures include soft tissue of the cheek(s) or forehead.
- hemifacial (nose, orbit and adjacent tissues) - restores part or all of the nose, the orbit, and significant adjacent facial tissue/structures, does not include any intraoral maxillary prosthesis or ocular prosthesis.
- partial facial prosthesis - restores a portion of the face, does not specifically involve the nose, orbit or ear
- nasal septal prosthesis - prosthesis that occludes a hole in the nasal septum, does not include superficial nasal tissue

Prosthetic devices may be secured or retained in place by anatomical structures; however, in most cases the device is held in place with the use of a skin adhesive. Additionally, some devices may be held in place by implants, such as bone integrated titanium implants. The method chosen to secure the device and the type of device are usually dependent upon factors such as the degree of deformity, the person's ability to handle maintenance routines, the individual's occupation and lifestyle, and the availability of assistance when needed.

Skin care products (e.g., cosmetics, creams, and cleansers) related to care of the prosthesis, and the application and/or removal of the device are considered personal care items.

UPPER LIMB: Myoelectric Prosthetic Device

The conventional prosthetic appliance for replacement of an upper extremity, either below or above the elbow, is a body-powered prosthesis with a terminal hand or hook device. A myoelectric device functions by means of electrical impulses and operates on rechargeable batteries requiring external cables or harnesses. It is a prosthetic device used as an alternative to a passive or conventional body-powered device which enables an amputee to adjust the force of his/her grip and an ability to both open and close the hand voluntarily. Myoelectric devices may be recommended for amputees who are unable to use body-powered devices or who require improved grip function/motion for performance of daily activities. Adults or children with above- or below-the-elbow amputations may use the device effectively, although as a child grows the prosthesis may require multiple socket replacements for proper fit and function.

A hybrid prosthesis is a device that uses a combination of myoelectric and body-powered technology to enhance the amputee's overall functionality, depending on the level and location of amputation. A hybrid device is indicated for high level amputations, (i.e, at or above the elbow) and consists of a body-powered device to control shoulder and elbow movement and a myoelectric device to control hand and wrist motion, allowing control of two joints at one time.

Literature Review

Results of studies published in the peer-reviewed scientific literature evaluating the impact of these devices on clinical outcomes are mixed (Egermann, et al., 2009; Pylatiuk, et al., 2007; Biddess and Chau, 2007; Crandall and Tomhave, 2002; Edelstein and Berger, 1993; Stein and Walley, 1983). Evidence is primarily in the form of small case series and does not provide strong conclusions to support the use of these devices for improving quality of life, although some authors have reported greater function and range of motion among subjects using the device (Crandall, Tomhave, 2002; Stein and Walley, 1983). Edelstein and Berger (1993) reported that activities such as donning socks, cutting paper and applying bandages were performed more rapidly with a myoelectric device when compared to a body-powered device, although performance with both devices was rated poorer than normal quality. In general, the reported outcomes are subjective and include patient acceptance and reasons for disuse; little data regarding functional status and direct comparisons to body-powered devices or passive devices are available. Patient selection criteria are not clearly defined, however despite these and other confounding variables, the published literature tends to support some clinical benefits from the use of a myoelectric prosthesis when compared to a conventional passive or body-powered device.

Areas of development for powered upper limb prosthetic devices include devices that function using implantable sensors, reinnervation of muscle fibers to allow fine movement control as well as sensory feedback and multiple simultaneous degrees of freedom. The LUKE (Life Under Kinetic Evolution) Arm (Mobius Bionics, LLC) is an upper limb prosthesis that has been developed to restore function in individuals who have lost all or part of their upper limb and that has multiple powered joints and grip patterns and is capable of multiple simultaneous degrees of freedom, controlled using EMG signals. In addition to the EMG electrodes, the LUKE Arm contains a combination of mechanisms, including switches, movement sensors, and force sensors. The primary control resides with inertial measurement sensors on top of the feet. The micro-electromechanical control system is operated through an inertial measurement unit (IMU), which is located in a sensor that is attached to or embedded in the individual's shoe. The user commands motion of the prosthesis by moving the foot in various directions. The device is available for transradial, transhumeral or shoulder amputation. Evidence in the peer reviewed literature is insufficient to support safety and efficacy of these emerging-type of devices. Within a Hayes report published in 2018 and updated in 2019, the authors concluded "the very-low-quality body of evidence is insufficient to draw conclusions regarding the efficacy and safety of the LUKE arm. The limited evidence does not suggest consistent improvement on measures of function or performance with the LUKE arm compared with existing prostheses and lacks adequate follow-up to compensate for the potential patient learning curve associated with the new prosthesis."(Hayes, 2019). Hayes noted the evidence reviewed included only two 2 pretest/posttest studies with internal controls; one was a Veterans' Affairs (VA) study to optimize the arm system (optimization study) and the second a VA home study of an advanced upper limb prosthesis.

LOWER LIMB PROSTHETIC DEVICE

Prior to being fitted with a lower limb prosthetic device, the individual must demonstrate specific functional levels. A functional level is defined as a measurement of the capacity and potential of the individual to accomplish his/her expected post-rehabilitation daily function.

Lower limb prosthetic devices may be preparatory or permanent. A preparatory device is a prosthesis made soon after an amputation (approximately four weeks) as a temporary method of retraining a person to walk and balance while shrinking the residual limb. A permanent prosthesis is recommended when an individual has used a prosthetic device full time for a period of six months and when the limb volume has stabilized to a point where the socket fit remains relatively consistent for 2–3 weeks.

Components and/or additions to a prosthesis may be medically necessary; the determination of medical necessity is based on the person's functional ability and expected functional potential as defined by the prosthetist and the ordering physician. Additional documentation supporting medical necessity must accompany claims submitted for prosthetic components and/or additions. Customizing prosthetic devices with enhanced features is not medically necessary if activities of daily living can be met with standard devices.

Accessories that are necessary for the effective use of the prosthetic device may also be considered medically necessary devices. Accessories that are not necessary for the effective use of the device are considered not medically necessary. While some prosthetic manufacturers offer devices with waterproof features, including devices that are submersible (e.g., Water Leg, [Standard Cyborg, SF, CA] [used for showering, swimming], Genium X3 [Ottobock, US], [a waterproof microprocessor-controlled knee prosthetic device]), when used for recreational purposes these prosthetic accessories/devices are considered a convenience item and not medically necessary.

Osseointegrated Prosthesis

Additionally, more advanced technological systems using multiple sensors to send messages back to a microchip regarding changes in walking patterns and osseanchored prosthetic devices for lower limbs are being investigated. These devices represent emerging technologies and are undergoing clinical trials evaluating performance, safety and durability. In contrast to the standard of care socket-suspended prosthesis, an osseanchored prosthetic device consists of a fixture and abutment screw that is surgically implanted into bone. After healing and various stages of rehabilitation the fixture is then attached to a prosthesis. One such device, the OPRA™ Implant System (Integrum AB, Sweden), has received FDA approval as a humanitarian device for prosthetic use. According to the manufacturer the system consists of three parts; an anchoring element (the Fixture) and a skin penetrating connection (the Abutment), and a securing titanium screw (the Abutment Screw). FDA labeling indicates the intended use is for patients who have transfemoral amputation due to trauma or cancer and who have rehabilitation problems with, or cannot use, a conventional socket prosthesis. The OPRA device is intended for skeletally mature patients. According to Hayes, evidence in the peer reviewed scientific literature evaluating this technology is insufficient to support safety and efficacy (Hayes, 2018). A systematic review published in 2018 by Kunutsor and colleagues evaluating the safety and efficacy of osseointegrated prostheses included a total of 22 eligible articles; 13 of the studies were unique. The average sample size of the studies included ranged from 11 to 100 participants, none of the studies were RCTs. The reported outcomes of all studies supported improvement in functional outcomes (walking ability, prosthetic use and mobility), and satisfaction and quality of life following osseointegration, compared with their preoperative status or when using a conventional socket prosthesis. Infection rates varied from 1% to 77%, with the majority of infections described as low-grade soft tissue or superficial infections related to the skin–implant interface. Infections were treated effectively with antibiotics. According to the authors none of the studies reported additional amputation or death as a result of osseointegration and they concluded osseointegration following limb amputation improved prosthetic use, comfort when sitting, walking ability, mobility, gait and quality of life. However, use of such devices is associated with an increased risk of soft tissue infection.

In 2017 the Canadian Agency for Drugs and Technologies in Health (CADTH) published a systematic review to evaluate the evidence for osseointegrated prosthetic devices for lower limb amputation. After reviewing seven studies that met inclusion criteria, the authors concluded the quality of evidence is generally low, and while some evidence suggests there is improvement in quality of life, function and mobility after implantation there is concern regarding high rates of infection, and the design of and materials used affecting safety and efficacy. Overall, the authors reported the available evidence suggests that careful attention should be given to patient

selection, implant selection, and residual limb skin integration, as well as surgical and rehabilitation protocols, to optimize outcomes and reduce adverse event rates.

LOWER LIMB: Microprocessor-controlled Device

Microprocessor-controlled Knee: Microprocessor-controlled knee prosthetics are sensor-equipped devices. The sensor detects when the knee is in full extension and adjusts the swing phases automatically, allowing a more natural pattern of walking at variable speeds (passive powered device). Multiple devices are available that use various degrees of computer technology to enhance the clinical function of the basic mechanical knee design; all microprocessor controlled systems do not have identical features and functions. Some devices have swing only, stance phase only, or swing and stance phase. Some of the devices currently available include but are not limited to the Otto Bock C-Leg[®], Genium and X2 (Otto Bock HealthCare, Minneapolis, MN), and the Endolite Orion, Intelligent and SmartIP (Endolite North America, Chase A. Blatchford and Sons Ltd., Miamisburg, OH). Another microprocessor device, the X3 (Otto Bock HealthCare, Minneapolis, MN), is waterproof; the device is completely submersible according to the manufacturer. A number of other devices are currently under investigation.

The purported advantages of a microprocessor controlled above-the-knee (AKA) prosthesis include:

- reduced energy expenditure of the amputee
- improved ability to walk on uneven ground
- improved ability to climb and descend stairs
- increased walking distance

Literature Review: In the published, peer-reviewed scientific literature, evidence supporting the use of microprocessor-controlled/computer-controlled prostheses comes primarily from small-group case studies with few randomized, case-controlled trials, and few systematic reviews. Of the groups studied clinically, most individuals were in good health and without other medical complications. Evidence in the peer-reviewed, published scientific literature does support reduction in energy consumption and a more symmetrical gait pattern when compared to a standard device (Datta and Howitt, 1998; Datta, et al., 2005; Seymour, et al., 2007; Highsmith, et al., 2010; Aldridge Whitehead, et al., 2014). Some evidence supports both reduced hip moment and metabolic requirements particularly at faster speeds. Although the evidence continues to evolve, there is evidence that supports the effective use of these devices for limited populations (Kaufmann, et al., 2007; Hafner, et al., 2007; Seymour, et al., 2007; Orenduff, et al., 2006; Datta, et al., 2005; Chin, et al., 2003; Datta and Howitt, 1998).

Microprocessor-controlled Ankle: In order to enhance the basic mechanical design and mimic the action of a biological ankle researchers have applied microprocessor technology to prosthetic feet (e.g., Proprio Foot, Ossur, élan Foot, Endolite). Stair ambulation is limited in the transtibial amputee as a result of neutral and fixed ankle position. Newer prosthetic ankles which adjust for ankle angle during swing phase and identify sloping gradients and ascent or descent of stairs are under investigation. One microprocessor-controlled ankle foot prosthesis currently available which has received FDA approval is the Proprio Foot[®] (Ossur, ALiso Viejo, CA). The Proprio Foot is a quasi-passive ankle that is able to actively change the ankle angle in the unloaded swing phase as the result of microprocessor-control and sensor technology. The device is passive (without power) while in stance phase. According to the manufacturer the proposed benefits of microprocessor-controlled ankle movements include the ability to identify slopes and stairs, when ascending or descending stairs the device automatically adapts ankle position to enable the next step; allows the user to place both feet behind their knees when rising from a chair; and automatically gives a toe-lift allowing sufficient ground clearance when walking. The device is designed to promote a more symmetrical and balanced gait and is intended for use by transtibial amputees engaging in low to moderate impact activities who are classified as level K3 (i.e., community ambulatory, with the ability or potential for ambulation with variable cadence); it is not suitable for sport and high impact activities.

Literature Review: Evidence in the published peer-reviewed scientific literature evaluating the use of microprocessor-controlled ankle foot devices is limited and consists mainly of pilot studies and case series involving small sample populations (Struchkov, Buckley, 2016; DeLussu, et al., 2013; Agrawal, et al., 2013; Darter and Wilkin, 2013; Fradet, et al., 2010; Alimusaj et al., 2009; Wolf, et al., 2009). Although limited, the

evidence does demonstrate some clinical advantages for use compared to conventional ankle foot prosthesis for individuals who are functional level 3 or greater.

Combination Microprocessor-controlled Knee-Ankle/Foot Prosthetic: Combination microprocessor prosthetics are available integrating both a microprocessor knee and the ankle/foot device (e.g., SYMBIONIC® LEG 3 [Ossur, Iceland]; LiNX® [Endolite]). One device, the SYMBIONIC® LEG 3 is a prosthetic that combines a microprocessor knee with a powered microprocessor ankle with proactive ankle flexion. The device purportedly has a more powerful knee actuator and new kinematic sensors for improved stability, increased support with stance flexion, and more rapid, and consistent swing extension. For a transfemoral amputee, combining both types of prosthetic devices theoretically enables a more natural and symmetrical gait when ambulating, decreasing energy expenditure, and offering increased stability. The device is intended for use by individuals who are Functional Level 3 or 4. The LiNX® [Endolite]) prosthetic system is intended for individuals who are Functional Level 3 or greater; according to the manufacturer this system is an integrated prosthetic utilizing a microprocessor-controlled system in addition to sensors and actuators which simultaneously controls the knee and foot.

Technology Assessment: The California Technology Assessment Forum (CTAF) conducted a review of the scientific evidence for the use of microprocessor-controlled prosthetic knees for individuals with trans-femoral amputations (CTAF, 2007). The forum recommended that use of C-Leg microprocessor-controlled prosthetic met CTAF criteria for safety, effectiveness, and improvement in health outcomes for a subset of individuals.

In a more recent report evaluating the C-Leg prosthesis for individuals with above-knee amputee, Hayes, Inc. concluded the available studies provided preliminary evidence that compared with mechanical prostheses the C-Leg prosthesis improves the mobility of selected unilateral transfemoral amputees (Hayes, 2015).

Professional Societies/Organizations: Although there are no recent updates, the United States Department of Veterans Affairs (VA, 2000), the Washington State Department of Labor and Industries (2002) and the Workers Compensation Board of British Columbia (2003, updated 2009) previously evaluated the clinical utility of microprocessor-controlled prosthetics. Within these publications the authors acknowledge, although not robust, there is some evidence to support a microprocessor knee improves function and/or is equivalent to a non-microprocessor device and supports use for some individuals.

LOWER LIMB: Powered Microprocessor-controlled Prosthetic Device

Powered Knee: Powered prosthetic devices that use signals from muscle activity in the remaining limb to bend and straighten the device remain under investigation. These devices utilize sensors and electronics to process data and control movement and power of the knee. Examples of this type of device include the Power Knee™, manufactured by Ossur (Foothill Ranch, CA). According to the manufacturer, the Power Knee is described as a motorized device which contains a rechargeable battery pack. It is designed to replace muscle activity of the quadriceps muscle and uses artificial proprioception with sensors in order to anticipate and respond with the appropriate movement required for stepping (active powered device). In comparison to a passive prosthetic knee, including a microprocessor device, the manufacturer suggests a power knee offers advantages such as powered extension with standing, controlled resistance with descending, and active flexion and extension during walking. The device controls the transition from a bent knee to an extended knee, at heel strike supports the individual's full body weight, and can help lift above-knee amputees out of a chair to a standing position. It is suggested the device helps to maintain walking speeds, assists with upward motion (required for stairs and inclines), and learns and responds to gait patterns. With the initial use of the device a practitioner must program and align the knee. Once programming and alignment are complete, the user needs only to press the power button to use the device. The device is compatible with a variety of dynamic flex-foot feet, must be re-charged daily and is not intended for high impact activity, sports, excessive loading or heavy duty use.

According to criteria outlined in the Centers for Medicare and Medicaid Services Local Coverage Determination, the following individuals may benefit from the use of a power knee-ankle device:

- the individual has a microprocessor (swing and stance phase type (L5856) controlled (electronic) knee
- is K3 functional level only
- has a documented comorbidity of the spine and/or sound limb affecting hip extension and/or quadriceps function that impairs K-3 level function with the use of a microprocessor-controlled knee alone

- is able to make use of a product that requires daily charging
- is able to understand and respond to error alerts and alarms indicating problems with the function of the unit.

Powered Foot-ankle: Similar to the powered knee device, powered foot-ankle prosthetic devices (HCPCS L5973 and L5969) are currently under investigation. Two such devices are the BiOM® Ankle and emPOWER™ Ankle, (BionX Medical Technologies, [previously iWalk, Inc., Bedford, MA). The BiOM device (previously referred to as Powerfoot One) uses a combination of processors, sensors, motors, and springs that allow the user a powered push-off with taking steps. Theoretically the device replaces the action of the foot, Achilles tendon and calf muscle to result in a near normalized gait for amputees and is intended for amputees that are functional level 3 or 4. According to the manufacturer, the emPOWER™ Ankle is a more recent generation of the BiOM® Ankle.

Literature Review: The available evidence in the published scientific literature consists mainly of preliminary studies evaluating device design and biomechanics with few comparative trials available. While some authors have reported on performance such as kinematic measures, improved energy costs, and biomechanical analysis (Simon, et al., 2016; Ingraham, et al., 2016; Gates, et al, 2013; Aldridge, et al., 2013) with the use of a powered prosthetic device (ankle/foot or knee), these studies involve small sample populations and evaluate short-term outcomes. Regarding the powered knee device, Wolfe et al. (2013) evaluated functional and clinical differences during sit-to-stand and step-up among power knee users (n=5) compared to the microprocessor C-Leg (n=5). The authors noted few differences between users during sit-to-stand and step-up task and no difference with regards to decreased impact on the intact limb. There is a paucity of published evidence evaluating ankle/foot powered devices (Rabago, et al., 2016; Esposito, et al., 2016; Takahashi, et al., 2013; Grabowski, DeAndrea, 2013; Herr, Grabowski, 2012). Until clinical trials are conducted to confirm the safety, efficacy and overall clinical utility of the powered ankle/foot device compared with other conventional or microprocessor prostheses, improvement in net health outcomes has yet to be determined.

Appendix 1 – Lower Limb Prosthetic “Device to Coding” Crosswalk

Please note, coding may vary according to manufacturer. This list is for informational purposes, it DOES NOT indicate coverage/non-coverage of a device.

Device Name	Brief Description	Manufacturer	Code(s)
Allux	Microprocessor-controlled knee	Nabtesco	L5613, L5845, L5848, L585, K1014
BiOM Foot	Microprocessor-controlled ankle foot (power)	BionX Medical Technologies	L5969, L5973
C-Leg	Microprocessor-controlled knee	Otto Bock	L5856, L5848, L5845, L5828
C-Leg Compact	Microprocessor-controlled knee	Otto Bock	L5858, L5828, L5848
Élan Foot	Microprocessor-controlled ankle foot	Endolite	L5973
EmPOWER™	Microprocessor-controlled ankle foot (power)	BionX Medical Technologies	L5969, L5973
Genium	Microprocessor-controlled knee	Otto Bock	L5999, L5999, L5999, L5848, L5828, L5850, and L5930
Genium X2, Genium X3	Microprocessor-controlled knee (X3 is water proof)	Otto Bock	L5999
IP +	Microprocessor-controlled knee	Endolite	L5857, L5830, L5616, L5845
Kinnex Foot	Microprocessor ankle/foot (waterproof)	Freedom Innovations	L5973

LiNX®	Combination microprocessor-controlled knee and foot; additionally has sensors and actuators	Endolite	L5856, L5848, L5845, L5828, L5830, L5973
Meridium Foot	Microprocessor-controlled ankle foot	Otto Bock	L5999
Orion	Microprocessor-controlled knee	Endolite	L5856, L5848, L5845, L5828, L5830
Össur Power Knee	Motor powered knee	Össur	L5859, L5856, L5828, L5848, L5845
Plié 3	Microprocessor-controlled knee (submersible)	Freedom Innovation (Freedom innovation component recently purchased by Otto Bock)	L5856, L5848, L5845, L5828
Proprio Foot®	Microprocessor-controlled ankle foot	Össur	L5973
Raize Foot	Microprocessor foot (does not have the power ankle)	Fillauer	L5973
Rheo	Microprocessor-controlled knee	Össur	L5856, L5848, L5845, L5828
Rheo XC	Microprocessor-controlled knee (supports rehabilitation to full recovery)	Össur	L5856, L5848, L5845, L5828
SYMBIONIC® LEG 3	Combination microprocessor-controlled knee and ankle with proactive ankle flexion	Össur	L5856, L5848, L5845, L5828, L5973

LOWER LIMB: Vacuum Suspension System

Suspension systems for lower limb prostheses keep the prosthesis in place, ensuring a good fit between the socket and residual limb. The intended function is to provide a connection that reduces rotational and shearing forces which can result in skin breakdown as well provide for balance and steady gait. Various types of suspension systems are available and include those that are primarily mechanical or suction-type systems. Mechanical systems involve the use of belts, straps, or sleeves, for example, to attach the device to the residual limb (L5666, L5670-L5672). Suction-type systems function by way of a negative pressure created between the socket and insert/liner. These devices can be passive (air escapes while donning via a one-way valve) or active (suction pump evacuates the air). Passive systems involve the use of a soft liner, a one-way valve and a donning sleeve. A liner is placed over the limb, the limb is placed in the socket and the force of one's body weight upon standing expels excess air through the valve creating a seal. With active suction devices the sleeve creates a seal around the edge of the socket and a pump and exhaust remove the excess air between the socket and the liner to ensure a secure fit.

Various vacuum suction-type devices (mechanical or electrical) are available and include the Vacuum-Assisted Socket System (VASS™) (Otto Bock Harmony Vacuum-Assisted Socket System, Otto Bock HealthCare; Minneapolis, MN), the eVAC® (Smith, Global), and the LimbLogic™ VS prosthetic vacuum suspension system (Mount Sterling, Ohio). Each device is a vacuum suction-type suspension system that manufacturers claim helps control volume fluctuation in the residual limbs of lower-extremity amputees, reduces forces to the limbs, and improves both suspension and proprioception without restricting vascular flow. Although patient selection criteria have not been firmly established, the device has been proposed for individuals with non-healing skin ulcerations located on the stump and/or when other socket systems have failed to provide a secure fit.

Evidence in the published, peer-reviewed scientific literature evaluating suspension systems, in particular vacuum suction-type suspension systems is limited. Much of the published literature is in the form of feasibility trials, case reports, and uncontrolled case series involving small populations. Reported outcomes are mixed, are short term, lack high statistical power and cannot be generalized. The results of one published randomized trial (Traballesi, et al., 2012) demonstrated that following a 12 week rehabilitation program VASS users had better clinical mobility compared to subjects using a conventional prosthesis with a standard suction socket. The authors reported that VASS users used their prosthesis more than the control group and that despite increased use, pain while using the VASS device did not differ significantly compared to the control group at various points of follow-up. The sample size of the trial involved only 20 subjects, three of whom dropped out of the study, and therefore generalization of results to larger populations cannot be made.

Although there is no recent update, the Washington State Department of Labor and Industries (2003) concluded, after an evaluation conducted by the technology assessment committee of the Otto Bock Vacuum Assisted Socket System (VASS) device, that the published literature does not substantially support the device's effectiveness for maintaining limb volume.

The published evidence does not provide strong support of clinical utility for this technology compared to conventional socket-suspension systems for the general population and clinical effectiveness has not been firmly established. The choice of a suspension system is determined by factors such as activity level, residual stump shape, age, and health status. There is some evidence in the published literature to support vacuum systems decrease limb volume fluctuations, can improve socket fit, reduce inside movement for some individuals, as well as improve comfort and satisfaction (Gholizadeh, et al., 2016). While additional long term studies and higher quality data would be helpful for evaluating an active suction-type vacuum suspension system, for individuals where other types of suspension systems have failed to provide a secure fit or are contraindicated, a vacuum suction-type suspension system may be considered an effective alternative.

Medicare Coverage Determinations

	Contractor	Policy Name/Number	Revision Effective Date
NCD	National	No NCD	
LCD	CGS Administrators	Lower Limb Prosthesis (LCD L33787)	1/1/2020
LCD	Noridian Healthcare Solutions	Lower Limb Prosthesis (LCD L33787)	1/1/2020
LCD	CGS Administrators	Facial Prosthesis (LCD L33738)	1/1/2020
LCD	Noridian Healthcare Solutions	Facial Prosthesis (LCD L33738)	1/1/2020

Note: Please review the current Medicare Policy for the most up-to-date information.

Coding/Billing Information

Note: 1) This list of codes may not be all-inclusive.

2) Deleted codes and codes which are not effective at the time the service is rendered may not be eligible for reimbursement.

IRIS PROSTHESIS

Experimental/ Investigational/ Unproven

CPT®* Codes	Description
0616T	Insertion of iris prosthesis, including suture fixation and repair or removal of iris, when performed; without removal of crystalline lens or intraocular lens, without insertion of intraocular lens
0617T	Insertion of iris prosthesis, including suture fixation and repair or removal of iris, when performed; with removal of crystalline lens and insertion of intraocular lens

CPT®* Codes	Description
0618T	Insertion of iris prosthesis, including suture fixation and repair or removal of iris, when performed; with secondary intraocular lens placement or intraocular lens exchange

HCPCS Codes	Description
C1839	Iris prosthesis

EXTERNAL FACIAL PROSTHESIS

Considered Medically Necessary when criteria in the applicable policy statements listed above are met and only when coverage is available under the plan for the specific device/component/item.

Nasal Prosthesis

Considered Medically Necessary when criteria in the applicable policy statements listed above are met:

CPT®* Codes	Description
21087	Impression and custom preparation; nasal prosthesis

HCPCS Codes	Description
L8040	Nasal prosthesis, provided by a non-physician
L8047	Nasal septal prosthesis, provided by a non-physician

Orbit Prosthesis

Considered Medically Necessary when criteria in the applicable policy statements listed above are met:

CPT®* Codes	Description
21077	Impression and custom preparation; orbital prosthesis

HCPCS Codes	Description
L8042	Orbital prosthesis, provided by a non-physician

Ear Prosthesis

Considered Medically Necessary when criteria in the applicable policy statements listed above are met:

CPT®* Codes	Description
21086	Impression and custom preparation; auricular prosthesis

HCPCS Codes	Description
L8045	Auricular prosthesis, provided by a non-physician

Facial Prosthesis

Considered Medically Necessary when criteria in the applicable policy statements listed above are met:

CPT®* Codes	Description
21088	Impression and custom preparation; facial prosthesis

HCPCS Codes	Description
L8041	Midfacial prosthesis, provided by a non-physician
L8043	Upper facial prosthesis, provided by a non-physician
L8044	Hemi-facial prosthesis, provided by a non-physician
L8046	Partial facial prosthesis, provided by a non-physician

Maxillofacial Prosthesis, External

Considered Medically Necessary when criteria in the applicable policy statements listed above are met:

HCPCS Codes	Description
L8048	Unspecified maxillofacial prosthesis, by report, provided by a non-physician

Considered not medically necessary when used to report non-covered consumable supplies outlined in the coverage policy:

HCPCS Codes	Description
A4364	Adhesive, liquid or equal, any type, per ounce
A4450	Tape, non-waterproof, per 18 square inches
A4452	Tape, waterproof, per 18 square inches
A4455	Adhesive remover or solvent (for tape, cement or other adhesive), per ounce
A4456	Adhesive remover, wipes, any type, each
L9900	Orthotic and prosthetic supply, accessory, and/or service component of another HCPCS "L" code

UPPER LIMB ADDITIONS/COMPONENTS

Additional Components/Features of Non Myoelectric Prosthetic Device:

Considered medically necessary when used to report a medically necessary component or addition to an upper limb prosthetic device in the absence of a specific code.

HCPCS Codes	Description
L6646	Upper extremity addition, shoulder joint, multipositional locking, flexion, adjustable abduction friction control, for use with body powered or external powered system
L6647	Upper extremity addition, shoulder lock mechanism, body powered actuator
L7499†	Upper extremity prosthesis, not otherwise specified

†Note: Covered when used to report a medically necessary component or addition to an upper limb prosthetic device in the absence of a specific code.

UPPER LIMB: MYOELECTRIC PROSTHETIC DEVICE

Considered Medically Necessary when criteria in the applicable policy statements listed above are met and only when coverage is available under the plan for the specific device/component/item:

HCPCS Codes	Description
L6026	Transcarpal/metacarpal or partial hand disarticulation prosthesis, external power, self-suspended, inner socket with removable forearm section, electrodes and cables, two batteries, charger, myoelectric control of terminal device, excludes terminal device(s)
L6611	Addition to upper extremity prosthesis, external powered, additional switch, any type
L6638	Upper extremity addition to prosthesis, electric locking feature, only for use with manually powered elbow
L6646	Upper extremity addition, shoulder joint, multipositional locking, flexion, adjustable abduction friction control, for use with body powered or external powered system
L6647	Upper extremity addition, shoulder lock mechanism, body powered actuator
L6648	Upper extremity addition, shoulder lock mechanism, external powered actuator
L6715	Terminal device, multiple articulating digit, includes motor(s), initial issue or replacement
L6880	Electric hand, switch, or myoelectric controlled, independently articulating digits, any grasp pattern or combination of grasp patterns, includes motor(s)
L6881	Automatic grasp feature, addition to upper limb electric prosthetic terminal device
L6882	Microprocessor control feature, addition to upper limb prosthetic terminal device
L6920	Wrist disarticulation, external power, self-suspended inner socket, removable forearm shell, Otto Bock or equal, switch, cables, two batteries and one charger, switch control of terminal device
L6925	Wrist disarticulation, external power, self-suspended inner socket, removable forearm shell, Otto Bock or equal, electrodes, cables, two batteries and one charger, myoelectronic control of terminal device
L6930	Below elbow, external power, self-suspended inner socket, removable forearm shell, Otto Bock or equal switch, cables, two batteries and one charger, switch control of terminal device
L6935	Below elbow, external power, self-suspended inner socket, removable forearm shell, Otto Bock or equal electrodes, cables, two batteries and one charger, myoelectronic control of terminal device
L6940	Elbow disarticulation, external power, molded inner socket, removable humeral shell, outside locking hinges, forearm, Otto Bock or equal switch, cables, two batteries and one charger, switch control of terminal device
L6945	Elbow disarticulation, external power, molded inner socket, removable humeral shell, outside locking hinges, forearm, Otto Bock or equal electrodes, cables, two batteries and one charger, myoelectronic control of terminal device
L6950	Above elbow, external power, molded inner socket, removable humeral shell, internal locking elbow, forearm, Otto Bock or equal switch, cables, two batteries and one charger, switch control of terminal device
L6955	Above elbow, external power, molded inner socket, removable humeral shell, internal locking elbow, forearm, Otto Bock or equal electrodes, cables, two batteries and one charger, myoelectronic control of terminal device
L6960	Shoulder disarticulation, external power, molded inner socket, removable shoulder shell, shoulder bulkhead, humeral section, mechanical elbow, forearm, Otto Bock or equal switch, cables, two batteries and one charger, switch control of terminal device
L6965	Shoulder disarticulation, external power, molded inner socket, removable shoulder shell, shoulder bulkhead, humeral section, mechanical elbow, forearm, Otto Bock or equal electrodes, cables, two batteries and one charger, myoelectronic control of terminal device
L6970	Interscapular-thoracic, external power, molded inner socket, removable shoulder shell, shoulder bulkhead, humeral section, mechanical elbow, forearm, Otto Bock or equal switch, cables, two batteries and one charger, switch control of terminal device
L6975	Interscapular-thoracic, external power, molded inner socket, removable shoulder shell, shoulder bulkhead, humeral section, mechanical elbow, forearm, Otto Bock or equal electrodes, cables, two batteries and one charger, myoelectronic control of terminal device
L7007	Electric hand, switch or myoelectric controlled, adult
L7008	Electric hand, switch or myoelectric controlled, pediatric
L7009	Electric hook, switch or myoelectric controlled, adult
L7040	Prehensile actuator, switch controlled

HCPCS Codes	Description
L7045	Electric hook, switch or myoelectric controlled, pediatric
L7170	Electronic elbow, Hosmer or equal, switch controlled
L7180	Electronic elbow, microprocessor sequential control of elbow and terminal device
L7181	Electronic elbow, microprocessor simultaneous control of elbow and terminal device
L7185	Electronic elbow, adolescent, Variety Village or equal, switch controlled
L7186	Electronic elbow, child, Variety Village or equal, switch controlled
L7190	Electronic elbow, adolescent, Variety Village or equal, myoelectronically controlled
L7191	Electronic elbow, child, Variety Village or equal, myoelectronically controlled
L7259	Electronic wrist rotator, any type
L7499†	Upper extremity prosthesis, not otherwise specified

†Note: Considered Medically Necessary when used to report components and/or additions to an upper limb prosthetic myoelectric device, if coverage for a myoelectric prosthetic device is available, and when medical necessity criteria are met.

Experimental/ Investigational/ Unproven when used to report an upper limb sensor and myoelectric controlled prosthetic device with simultaneous multiple degrees of freedom (e.g., LUKE [Life Under Kinetic Evolution] Arm)

HCPCS Codes	Description
L7499	Upper extremity prosthesis, not otherwise specified

LOWER LIMB: MECHANICAL (NON-POWERED, NON MICROPROCESSOR)

Considered medically necessary when used to report a component or addition to a lower limb prosthetic device when criteria in the applicable policy statements listed above are met and when coverage is available under the plan for the specific device/component/item:

HCPCS Codes	Description
L5828†	Addition, endoskeletal knee-shin system, single axis, fluid swing and stance phase control
L5845††	Addition, endoskeletal knee-shin system, stance flexion feature, adjustable
L5848†	Addition to endoskeletal, knee-shin system, fluid stance extension, dampening feature, with or without adjustability
L5930†††	Addition, endoskeletal system, high activity knee control frame
L5981†	All lower extremity prostheses, flex-walk system or equal
L5999††††	Lower extremity prosthesis, not otherwise specified

†Note: Considered medically necessary for functional level 3 or above when medical necessity criteria has been met for the base device.

††Note: Considered medically necessary for functional level 1 or above when medical necessity criteria has been met for the base device.

†††Note: Requires K-4 functional level and when medical necessity criteria has been met for the base device.

††††Note: Considered medically necessary when used to report a medically necessary component or addition to a lower limb prosthetic device in the absence of a more specific code and when medical necessity criteria has been met for the base device.

LOWER LIMB MICROPROCESSOR-CONTROLLED PROSTHETIC DEVICES

Considered Medically Necessary when criteria in the applicable policy statements listed above are met and when benefits are available under the plan for a microprocessor-controlled prosthetic:

HCPSC Codes	Description
L5856	Addition to lower extremity prosthesis, endoskeletal knee-shin system, microprocessor control feature, swing and stance phase, includes electronic sensor(s), any type
L5857	Addition to lower extremity prosthesis, endoskeletal knee-shin system, microprocessor control feature, swing phase only, includes electronic sensor(s), any type
L5858	Addition to lower extremity prosthesis, endoskeletal knee shin system, microprocessor control feature, stance phase only, includes electronic sensor(s), any type
L5973	Endoskeletal ankle foot system, microprocessor controlled feature, dorsiflexion and/or plantar flexion control, includes power source

Additional Components/Features of Microprocessor-Controlled Prosthetic Devices:

Considered Medically Necessary when criteria in the applicable policy statements listed above are met and when benefits are available under the plan for a microprocessor-controlled prosthetic:

HCPSC Codes	Description
K1014	Addition, endoskeletal knee-shin system, 4 bar linkage or multiaxial, fluid swing and stance phase control
L5828	Addition, endoskeletal knee-shin system, single axis, fluid swing and stance phase control
L5845	Addition, endoskeletal knee-shin system, stance flexion feature, adjustable
L5848	Addition to endoskeletal, knee-shin system, fluid stance extension, dampening feature, with or without adjustability
L5920	Addition, endoskeletal system, above knee or hip disarticulation, alignable system
L5925	Addition, endoskeletal system, above knee, knee disarticulation or hip disarticulation, manual lock
L5930†	Addition, endoskeletal system, high activity knee control frame
L5950	Addition, endoskeletal system, above knee, ultra-light material (titanium, carbon fiber or equal)
L5999††	Lower extremity prosthesis, not otherwise specified

†Note: L5930 requires K-4 functional level.

††Note: Covered when used to report a medically necessary component/feature or addition to a lower limb prosthetic microprocessor-controlled device in the absence of a specific code.

LOWER LIMB: POWERED MICROPROCESSOR-CONTROLLED PROSTHETIC DEVICES

Considered Medically Necessary and when benefits are available for a power-controlled or power-assisted lower limb knee device (e.g., Ossur Power Knee):

HCPSC Codes	Description
L5859†	Addition to lower extremity prosthesis, endoskeletal knee-shin system, powered and programmable flexion/extension assist control, includes any type motor(s)

†Note: L5859 requires K-3 functional level; the device is not intended for high impact activity, sports, excessive loading or heavy duty use.

Microprocessor-Controlled Ankle Foot Prosthetic with Power Assist (e.g., BiOM® Ankle, emPOWER™ Ankle)

Experimental/Investigational/Unproven:

HCPCS Codes	Description
L5969	Addition, endoskeletal ankle-foot or ankle system, power assist, includes any type motor(s)
L5973	Endoskeletal ankle foot system, microprocessor controlled feature, dorsiflexion and/or plantar flexion control, includes power source

Additional Components/Features of Powered Prosthetic Devices, Including Power Assist Features:

Experimental/Investigational/Unproven when reported in addition to a non-covered power-controlled (L5859, L5973) or power-assisted (L5969) prosthetic device:

HCPCS Codes	Description
L5828	Addition, endoskeletal knee-shin system, single axis, fluid swing and stance phase control
L5845	Addition, endoskeletal knee-shin system, stance flexion feature, adjustable
L5848	Addition to endoskeletal knee-shin system, fluid stance extension, dampening feature, with or without adjustability
L5856	Addition to lower extremity prosthesis, endoskeletal knee-shin system, microprocessor control feature, swing and stance phase, includes electronic sensor(s), any type
L5969	Addition, endoskeletal ankle-foot or ankle system, power assist, includes any type motor(s)

LOWER LIMB: VACUUM SUSPENSION SYSTEM

Considered Medically Necessary when criteria in the applicable policy statements listed above are met:

HCPCS Codes	Description
L5781	Addition to lower limb prosthesis, vacuum pump, residual limb volume management and moisture evacuation system
L5782	Addition to lower limb prosthesis, vacuum pump, residual limb volume management and moisture evacuation system, heavy duty

REPAIR AND REPLACEMENT

Considered Medically Necessary when criteria in the applicable policy statements listed above are met:

HCPCS Codes	Description
L7510	Repair of prosthetic device, repair or replace minor parts
L7520	Repair prosthetic device, labor component, per 15 minutes
L8049	Repair or modification of maxillofacial prosthesis, labor component, 15 minute increments, provided by a non-physician

***Current Procedural Terminology (CPT®) ©2020 American Medical Association: Chicago, IL.**

References

1. Agrawal V, Gailey RS, Gaunaud IA, O'Toole C, Finnieston AA. Comparison between microprocessor-controlled ankle/foot and conventional prosthetic feet during stair negotiation in people with unilateral transtibial amputation. J Rehabil Res Dev. 2013;50(7):941-50.

2. Aldridge JM, Sturdy JT, Wilken JM. Stair ascent kinematics and kinetics with a powered lower leg system following transtibial amputation. *Gait Posture*. 2012 Jun;36(2):291-5.
3. Aldridge Whitehead JM1, Wolf EJ, Scoville CR, Wilken JM. Does a Microprocessor-controlled Prosthetic Knee Affect Stair Ascent Strategies in Persons With Transfemoral Amputation? *Clin Orthop Relat Res*. 2014 Feb 11.
4. Alimusaj M, Fradet L, Braatz F, Gerner HJ, Wolf SI. Kinematics and kinetics with an adaptive ankle foot system during stair ambulation of transtibial amputees. *Gait Posture*. 2009 Oct;30(3):356-63.
5. Beil TL, Street GM, Covey SJ. Interface pressures during ambulation using suction and vacuum-assisted prosthetic sockets. *J Rehabil Res Dev*. 2002 Nov-Dec;39(6):693-700
6. Bellmann M, Schmalz T, Blumentritt S. Comparative biomechanical analysis of current microprocessor-controlled prosthetic knee joints. *Arch Phys Med Rehabil*. 2010 Apr;91(4):644-52.
7. Bellmann M, Schmalz T, Ludwigs E et al. Immediate effects of a new microprocessor-controlled prosthetic knee joint: a comparative biomechanical evaluation. *Arch Phys Med Rehabil* 2012; 93(3):541-9.
8. Bergman K, Ornholmer L, Zackrisson K, Thyberg M. Functional benefit of an adaptive myoelectric prosthetic hand compared to a conventional myoelectric hand. *Prosthet Orthot Int*. 1992 Apr;16(1):32-7.
9. Biddiss EA, Chau TT. Upper limb prosthesis use and abandonment: a survey of the last 25 years. *Prosthet Orthot Int*. 2007 Sep;31(3):236-57.
10. Blatchford. Prosthetic Products. Knees. Accessed December 2, 2019. Available at URL address: <https://www.blatchfordus.com/products/category/knees/>
11. Bonnet C, Miller KM. Safety and efficacy of custom foldable silicone artificial iris implantation: prospective compassionate-use case series. *J Cataract Refract Surg*. 2020 Jun;46(6):893-901.
12. Bouwsema H, van der Sluis CK, Bongers RM. Learning to control opening and closing a myoelectric hand. *Arch Phys Med Rehabil*. 2010 Sep;91(9):1442-6.
13. Brunelli S, Delussu AS, Paradisi F, Pellegrini R, Trallesi M. A comparison between the suction suspension system and the hypobaric Iceross Seal-In® X5 in transtibial amputees. *Prosthet Orthot Int*. 2013 Dec;37(6):436-44.
14. California Technology Assessment Forum (CTAF). Microprocessor-controlled prosthetic knees. October 17, 2007. Accessed December 2, 2019. Available at URL address: http://icer-review.org/wp-content/uploads/2016/01/777_file_Microprocessor_W.pdf
15. Canadian Agency for Drugs and Technologies in Health (CADTH). Osseointegrated Prosthetic Implants for Lower Limb Amputation: A Review of Clinical Effectiveness, Cost Effectiveness and Guidelines. Feb 27, 2017. Accessed December 9, 2019. Available at URL address: <https://www.cadth.ca/sites/default/files/pdf/htis/2017/RC0856%20Osseointegrated%20Prosthetic%20Implants%20Final.pdf>
16. Carey SL, Lura DJ, Highsmith MJ; CP; FAAOP. Differences in myoelectric and body-powered upper-limb prostheses: Systematic literature review. *J Rehabil Res Dev*. 2015;52(3):247-62.
17. Centers for Medicare & Medicaid Services (CMS). Local Coverage Article: Facial Prostheses (A52463). Revision effective date: 3/1/2019. Accessed December 2, 2019. Available at URL address: <https://www.cms.gov/medicare-coverage-database/details/article-details.aspx?articleId=52463&ContrID=140>

18. Centers for Medicare & Medicaid Services (CMS). Local Coverage Article: Eye Prostheses (A52462). Revision effective date 3/1/2019. Accessed February 22, 2018. Available at URL address: <https://www.cms.gov/medicare-coverage-database/details/article-details.aspx?articleId=52462&ContrID=140>
19. Centers for Medicare and Medicaid Services (CMS) Local Coverage Determination. Lower Limb Prosthesis (L33787). Revision effective date 10/1/2018. Available at URL address: <https://www.cms.gov/medicare-coverage-database/details/lcd-details.aspx?LCDId=33787&ver=17&DocType=1&bc=AAIAAAAAAAAA&>
20. Cheesborough JE, Smith LH, Kuiken TA, Dumanian GA. Targeted muscle reinnervation and advanced prosthetic arms. *Semin Plast Surg.* 2015 Feb;29(1):62-72.
21. Chin T, Sawamura S, Shiba R, Oyabu H, Magakura Y, Takase I, et al. Effect of an intelligent prosthesis (IP) on the walking ability of young transfemoral amputees. Comparison of IP users with able-bodied people. *Am J Phys Med Rehabil.* 2003 Jun;82(6):447-51.
22. Collins, SH, Kuo, AD. Recycling energy to restore impaired ankle function during human walking. *PLoS One.* 2010;5(2):e9307.
23. Crandall RC, Tomhave W. Pediatric unilateral below-elbow amputees: retrospective analysis of 34 patients given multiple prosthetic options. *J Pediatr Orthop.* 2002 May 1;22(3):380-3.
24. Cummings DR, Kapp SL. Lower-limb pediatric prosthetics: general considerations and philosophy. American Academy of Orthotists and Prosthetists (AAOP). *Journal of Prosthetics and Orthotics.* 1992; 4(4):196-206.
25. D'Andrea S, Wilhelm N, Silverman AK, Grabowski AM. Does Use of a Powered Ankle-foot Prosthesis Restore Whole-body Angular Momentum During Walking at Different Speeds? *Clin Orthop Relat Res.* 2014 Apr 30.
26. Darter BJ, Sinitski K, Wilken JM. Axial bone-socket displacement for persons with a traumatic transtibial amputation: The effect of elevated vacuum suspension at progressive body-weight loads. *Prosthet Orthot Int.* 2015 Sep 30.
27. Darter BJ, Wilken JM. Energetic consequences of using a prosthesis with adaptive ankle motion during slope walking in persons with a transtibial amputation. *Prosthet Orthot Int.* 2013 Mar 22.
28. Datta D, Heller B, Howitt J. A comparative evaluation of oxygen consumption and gait pattern in amputees using Intelligent Prostheses and conventional damped knee swing-phase control. *Clin Rehabil.* 2005 Jun;19(4):398-403.
29. Datta D, Howitt J. Conventional versus microchip controlled pneumatic swing phase control for transfemoral amputees: user's verdict. *Prosthet Orthot Int.* 1998 Aug;22(2):129-35.
30. Datta D, Ibbotson V. Powered prosthetic hands in very young children. *Prosthet Orthot Int.* 1998 Aug;22(2):150-4.
31. Delussu AS, Brunelli S, Paradisi F, Iosa M, Pellegrini R, Zenardi D, Trallesi M. Assessment of the effects of carbon fiber and bionic foot during overground and treadmill walking in transtibial amputees. *Gait Posture.* 2013 Sep;38(4):876-82.
32. Edeer D, Martin CW. Evidence based practice group. British Columbia (BC). WorkSafe BC. UPDATE
33. Edelstein JE, Berger N. Performance comparison among children fitted with myoelectric and body-powered hands. *Arch Phys Med Rehabil.* 1993 Apr;74(4):376-80.

34. Edwards SG, Gellman H. Elbow and Above-elbow Amputations. Medscape. December 23, 2015. Updated Oct 2, 2019. Accessed December 2, 2019. Available at URL address: <http://emedicine.medscape.com/article/1237886-overview>
35. Egermann M, Kasten P, Thomsen M. Myoelectric hand prostheses in very young children. *Int Orthop*. 2009 Aug;33(4):1101-5.
36. Esquenazi A1, Talaty M, Packel A, Saulino M. The ReWalk powered exoskeleton to restore ambulatory function to individuals with thoracic-level motor-complete spinal cord injury. *Am J Phys Med Rehabil*. 2012 Nov;91(11):911-21.
37. Ferris AE, Aldridge JM, Rábago CA, Wilken JM. Evaluation of a Powered Ankle-Foot Prosthetic System During Walking. *Arch Phys Med Rehabil*. 2012 Jun 22.
38. Figueiredo GB, Snyder ME. Long-term follow-up of a custom-made prosthetic iris device in patients with congenital aniridia. *J Cataract Refract Surg*. 2020 Jun;46(6):879-887.
39. Fineberg DB1, Asselin P, Harel NY, Agranova-Breyter I, Kornfeld SD, Bauman WA, Spungen AM. Vertical ground reaction force-based analysis of powered exoskeleton-assisted walking in persons with motor-complete paraplegia. *J Spinal Cord Med*. 2013 Jul;36(4):313-21.
40. Fradet L, Alimusaj M, Braatz F, Wolf SI. Biomechanical analysis of ramp ambulation of transtibial amputees with an adaptive ankle foot system. *Gait Posture*. 2010 Jun;32(2):191-8.
41. Frisina R, De Biasi CS, Londei D, et al. A new intraocular lens with artificial iris for treating a case of iris extrusion secondary to traumatic opening of a radial keratotomy. *Eur J Ophthalmol*. 2020 Jan 23.
42. Fuenzalida Squella SA, Kannenberg A, Brandão Benetti Â. Enhancement of a prosthetic knee with a microprocessor-controlled gait phase switch reduces falls and improves balance confidence and gait speed in community ambulators with unilateral transfemoral amputation. *Prosthet Orthot Int*. 2017 Jul 1:309364617716207.
43. Gailey RS, Gaunaured I, Agrawal V, Finnieston A, O'Toole C, Tolchin R. Application of self-report and performance-based outcome measures to determine functional differences between four categories of prosthetic feet. *J Rehabil Res Dev*. 2012 Jun;49(4):597-612.
44. Gailey RS, Gaunaured I, Agrawal V, Finnieston A, O'Toole C, Tolchin R. Application of self-report and performance-based outcome measures to determine functional differences between four categories of prosthetic feet. *J Rehabil Res Dev*. 2012;49(4):597-612.
45. Gates DH, Aldridge JM, Wilken JM. Kinematic comparison of walking on uneven ground using powered and unpowered prostheses. *Clin Biomech (Bristol, Avon)*. 2013 Apr;28(4):467-72.
46. Gerschutz M, Denune JA, Colvin JM, Schober G. Elevated Vacuum Suspension Influence on Lower Limb Amputee's Residual Limb Volume at Different Vacuum Pressure Settings. *JPO Journal of Prosthetics & Orthotics*. 2010 Oct; 22(4):252-256.
47. Gholizadeh H, Abu Osman NA, Eshraghi A, Ali S. The effects of suction and pin/lock suspension systems on transtibial amputees' gait performance. *PLoS One*. 2014 May 14;9(5):e94520.
48. Gholizadeh H, Lemaire ED, Eshraghi A. The evidence-base for elevated vacuum in lower limb prosthetics: Literature review and professional feedback. *Clin Biomech (Bristol, Avon)*. 2016 Jun 22;37:108-116.
49. Goswami J, Lynn R, Street G, Harlander M. Walking in a vacuum-assisted socket shifts the stump fluid balance. *Prosthet Orthot Int*. 2003 Aug;27(2):107-13.

50. Grabowski AM, D'Andrea S. Effects of powered ankle-foot prosthesis on kinetic loading of the unaffected leg during level-ground walking. *Neuroeng Rehabil*. 2013 Jun 7;10:49.
51. Hafner BJ, Willingham LL, Buell NC, Allyn KJ, Smith DG. Evaluation of function, performance, and preference as transfemoral amputees transition from mechanical to microprocessor control of the prosthetic knee. *Arch Phys Med Rehabil*. 2007 Feb;88(2):207-17.
52. Hayes, Clinical Research Response. Osseointegrated Prostheses for the Rehabilitation of Amputees (OPRA) (Integrum AB). Lansdale, PA: Hayes, Inc. Aug 8, 2018. Archived Sep 2019.
53. Hayes, Inc. Health Technology Assessment. The LUKE Arm (Mobius Bionics LLC) for Upper Extremity Amputation. Lansdale, PA: Hayes, Inc. May 10, 2018, updated June 3, 2019.
54. Hayes, Inc. Search and Summary. CircumFlex™ Artificial Iris. Lansdale, PA: Hayes, Inc. Jan 18 2019.
55. Hayes Inc., Search and Summary. Electric and Body-Powered Fingers (Advanced Arm Dynamics) for Partial Hand Replacement. Lansdale, PA: Hayes, Inc. November, 2014.
56. Hayes Inc., Search and Summary. Myoelectric Prosthetic Devices for Hand Replacement Following Amputation. Lansdale, PA: Hayes, Inc. August 2014.
57. Hayes, Inc. Health Technology Brief. The available studies provide preliminary evidence that, compared with mechanical prostheses, the C-Leg prosthesis improves the mobility of selected unilateral transfemoral amputees. Lansdale, PA: Hayes, Inc. Published Jan 30, 2013; reviewed Jan 12, 2015, archived March 2016.
58. Herr HM, Grabowski AM. Bionic ankle-foot prosthesis normalizes walking gait for persons with leg amputation. *Proc Biol Sci*. 2012 Feb 7;279(1728):457-64.
59. Highsmith MJ, Kahle JT, Bongiorno DR, Sutton BS, Groer S, Kaufman KR. Safety, energy efficiency, and cost efficacy of the C-Leg for transfemoral amputees: A review of the literature. *Prosthet Orthot Int*. 2010 Dec;34(4):362-77.
60. Highsmith MJ, Kahle JT, Miro RM, Mengelkoch LJ. Ramp descent performance with the C-Leg and interrater reliability of the Hill Assessment Index. *Prosthet Orthot Int* 2013 Jan 17.
61. Hijjawi JB, Kuiken TA, Lipschutz RD, Miller LA, Stubblefield KA, Dumanian GA. Improved myoelectric prosthesis control accomplished using multiple nerve transfers. *Plast Reconstr Surg*. 2006 Dec;118(7):1573-8.
62. Hofstad C, Linde H, Limbeek J, Postema K. Prescription of prosthetic ankle-foot mechanisms after lower limb amputation. *Cochrane Database Syst Rev*. 2004;(1):CD003978. Copyright © 2009 The Cochrane Collaboration.
63. Hoskins RD1, Sutton EE, Kinor D, Schaeffer JM, Fatone S. Using vacuum-assisted suspension to manage residual limb wounds in persons with transtibial amputation: a case series. *Prosthet Orthot Int*. 2014 Feb;38(1):68-74.
64. Ingraham KA, Fey NP, Simon AM, Hargrove LJ. Assessing the Relative Contributions of Active Ankle and Knee Assistance to the Walking Mechanics of Transfemoral Amputees Using a Powered Prosthesis. *PLoS One*. 2016 Jan 25;11(1):e0147661.
65. Jhamtani R, Meenakshi A, Thulasigam C, Sabarigirinathan C, Sahu B. Electro-Mechanical Finger Prosthesis: A Novel Approach for Rehabilitation of Finger Amputees. *J Indian Prosthodont Soc*. 2014 Dec;14(Suppl 1):150-4.

66. Kahle JT, Highsmith MJ. Transfemoral sockets with vacuum-assisted suspension comparison of hip kinematics, socket position, contact pressure, and preference: ischial containment versus brimless. *J Rehabil Res Dev*. 2013;50(9):1241-52.
67. Kaufman KR, Levine JA, Brey RH, Iverson BK, McCrady SK, Padgett DJ, Joyner MJ. Gait and balance of transfemoral amputees using passive mechanical and microprocessor-controlled prosthetic knees. *Gait Posture*. 2007 Oct;26(4):489-93.
68. Kaufman KR, Levine JA, Brey RH, McCrady SK, Padgett DJ, Joyner MJ. Energy expenditure and activity of transfemoral amputees using mechanical and microprocessor-controlled prosthetic knees. *Arch Phys Med Rehabil*. 2008 Jul;89(7):1380-5.
69. Klute GK, Berge JS, Biggs W, Pongnumkul S, Popovic Z, Curless B. Vacuum-assisted socket suspension compared with pin suspension for lower extremity amputees: effect on fit, activity, and limb volume. *Arch Phys Med Rehabil*. 2011 Oct;92(10):1570-5.
70. Komolafe O, Wood S, Caldwell R, Hansen A, Fatone S. Methods for characterization of mechanical and electrical prosthetic vacuum pumps. *J Rehabil Res Dev*. 2013;50(8):1069-78.
71. Kuiken T. Targeted reinnervation for improved prosthetic function. *Phys Med Rehabil Clin N Am*. 2006 Feb;17(1):1-13.
72. Kuiken TA, Dumanian GA, Lipschutz RD, Miller LA, Stubblefield KA. The use of targeted muscle reinnervation for improved myoelectric prosthesis control in a bilateral shoulder disarticulation amputee. *Prosthet Orthot Int*. 2004 Dec;28(3):245-53.
73. Kuiken TA, Li G, Lock BA, Lipschutz RD, Miller LA, Stubblefield KA, Englehart KB. Targeted muscle reinnervation for real-time myoelectric control of multifunction artificial arms. *JAMA*. 2009 Feb 11;301(6):619-28.
74. Kuiken TA, Marasco PD, Lock BA, Harden RN, Dewald JP. Redirection of cutaneous sensation from the hand to the chest skin of human amputees with targeted reinnervation. *Proc Natl Acad Sci USA*. 2007 Dec;104(50):20061-6.
75. Kuntze Ferreira AE, Neves EB. A comparison of vacuum and KBM prosthetic fitting for unilateral transtibial amputees using the Gait Profile Score. *Gait Posture*. 2015;41(2):683-687.
76. Kunutsor SK, Gillatt D, Blom AW. Systematic review of the safety and efficacy of osseointegration prosthesis after limb amputation. *Br J Surg*. 2018 Dec;105(13):1731-1741.
77. Laferrier JZ, Gailey R. Advances in lower-limb prosthetic technology. *Phys Med Rehabil Clin N Am*. 2010 Feb;21(1):87-110.
78. Lake C, Miguelez JM. Comparative analysis of microprocessors in upper limb prosthetics. *Journal of Prosthetics and Orthotics*. 2003;15(2):48-65.
79. Ledoux ED, Lawson BE, Shultz AH, Bartlett HL, Goldfarb M. Metabolics of stair ascent with a powered transfemoral prosthesis. *Conf Proc IEEE Eng Med Biol Soc*. 2015;2015:5307-10.
80. Liberating Technologies. LTI Boston Digital™ Arm Systems for Adults. Accessed December 2, 2019. Available at URL address: <http://www.liberatingtech.com/>
81. Mâaref K, Martinet N, Grumillier C, Ghannouchi S, André JM, Paysant J. Kinematics in the terminal swing phase of unilateral transfemoral amputees: microprocessor-controlled versus swing-phase control prosthetic knees. *Arch Phys Med Rehabil*. 2010 Jun;91(6):919-25.

82. Mancinelli, C, Patriitti, BL, Tropea, P, Greenwald, RM, Casler, R, Herr, H, and Bonato, P. Comparing a passive-elastic and a powered prosthesis in transtibial amputees. *Conf Proc IEEE Eng Med Biol Soc.* 2011;2011:8255-8258.
83. Maruishi M, Tanaka Y, Muranaka H, Tsuji T, Ozawa Y, Imaizumi S, et al. Brain activation during manipulation of the myoelectric prosthetic hand: a functional magnetic resonance imaging study. *Neuroimage.* 2004 Apr;21(4):1604-11.
84. Mayer CS, Laubichler AE, Khoramnia R, et al. Challenges and complication management in novel artificial iris Implantation. *J Ophthalmol.* 2018 Sep 23; 2018:3262068.
85. Mayer CS, Laubichler AE, Masyk M, et al. Residual iris retraction syndrome after artificial iris implantation. *Am J Ophthalmol.* 2019 Mar;199:159-166.
86. Mayer CS, Reznicek L, Hoffmann AE. Pupillary reconstruction and outcome after artificial iris implantation. *Ophthalmology.* 2016 May; 123(5):1011-1018.
87. Michaud LJ. Upper extremity prostheses. 6.3.5 Prostheses. Therapy services and anticipatory guidance for the child with disability. Chapter 6. Childhood disability and rehabilitation. In: Rudolph CD, Rudolph AM, Hostetter MK, Lister GE, Siegel NJ, editors. *Rudolph's pediatrics.* 21st ed. New York, NY: McGraw-Hill; 2003.
88. Mostafa YS, Osman AA, Hassanein DH, et al. Iris reconstruction using artificial iris prosthesis for management of aniridia. *Eur J Ophthalmol.* 2018 Jan;28(1):103-107.
89. Motion Control, Inc. Utah Arm 3. Accessed December 2, 2019. Available at URL address: <http://www.utaharm.com/ua3-myoelectric-arm.php>
90. National Institute for Health and Care Excellence (NICE). Direct skeletal fixation of limb or digit prostheses using intraosseous transcutaneous implants. IPG 270. Published July 2008. Accessed December 9, 2019. Available at URL address: <https://www.nice.org.uk/guidance/ipg270/chapter/1-Guidance>.
91. National Institute for Health and Care Excellence (NICE). Artificial iris insertion for acquired aniridia. Interventional procedures guidance (IPG)674. Published July 22, 2020. Accessed December 8, 2020. Available at URL address: <https://www.nice.org>
92. Orendurff MS, Segal AD, Klute GK, McDowell ML, Pecoraro JA, Czerniecki JM. Gait efficiency using the C-Leg. *J Rehabil Res Dev.* 2006 Mar-Apr;43(2):239-46.
93. Ossur, UA. Power knee. Accessed December 2, 2019. Available at URL address: <https://www.ossur.com/prosthetic-solutions/products/dynamic-solutions/power-knee>
94. Otr OV, Reinders-Messelink HA, Bongers RM, Bouwsema H, Van Der Sluis CK. The i-LIMB hand and the DMC plus hand compared: a case report. *Prosthet Orthot Int.* 2010 Jun;34(2):216-20.
95. Otto Bock C-Leg® and more: A review of microprocessor-controlled prosthetic knees. January 2009. Accessed July 30, 2015. Available at URL address: http://worksafebc.com/health_care_providers/Assets/PDF/Otto_Bock_CLeg_2009_Update.pdf
96. Otto Bock®. C-Leg® System. December 2, 2019. Available at URL address: http://www.ottobockus.com/prosthetics/lower-limb-prosthetics/?utm_source=google&utm_medium=cpc&utm_term=microprocessor%20leg&utm_campaign=US%20Microprocessor%20Knee%20General

97. Pasquina PF, Evangelista M, Carvalho AJ, et al. First-in-man demonstration of a fully implanted myoelectric sensors system to control an advanced electromechanical prosthetic hand. *J Neurosci Methods*. 2014 Aug 4. pii: S0165-0270(14)00267-2. Epub.
98. Pinzur MS. Prosthetic considerations. In: Browner BD, Jupiter JB, Levine AM, Trafton PG, editors. *Skeletal Trauma: Basic Science, Management, and Reconstruction*, 3rd ed. 2003. Chapter 64.
99. Pomeranz B, Adler U, Shenoy N, Macaluso C, Parikh S. Prosthetics and orthotics for the older adult with a physical disability. *Clin Geriatr Med*. 2006 May;22(2):377-94;ix.
100. Prinsen EC, Nederhand MJ, Sveinsdóttir HS, et al. The influence of a user-adaptive prosthetic knee across varying walking speeds: A randomized cross-over trial. *Gait Posture*. 2017 Jan;51:254-260. doi: 10.1016/j.gaitpost.2016.11.015.
101. Pylatiuk C, Schulz S, Döderlein L. Results of an Internet survey of myoelectric prosthetic hand users. *Prosthet Orthot Int*. 2007 Dec;31(4):362-70.
102. Rábago CA, Aldridge Whitehead J, Wilken JM. Evaluation of a Powered Ankle-Foot Prosthesis during Slope Ascent Gait. *PLoS One*. 2016 Dec 15;11(12):e0166815.
103. Rickmann A, Szurman P, Januschowski K, et al. Long-term results after artificial iris implantation in patients with aniridia. *Graefes Arch Clin Exp Ophthalmol*. 2016 Jun; 254(7):1419-1424.
104. Rossbach, P. Amputee Coalition of America (ACA). When to replace a prosthesis. Fact sheet. National Limb Loss Information Center. ©Copyright Amputee Coalition. Reviewed and updated 01/2008. Accessed December 2, 2019. Available at URL address: <http://www.amputee-coalition.org/resources/when-to-replace-a-prosthesis/>
105. Routhier F, Vincent C, Morissette MJ, Desaulniers L. Clinical results of an investigation of paediatric upper limb myoelectric prosthesis fitting at the Quebec Rehabilitation Institute. *Prosthet Orthot Int*. 2001 Aug 1;25(2):119-31.
106. Russell Esposito E, Aldridge Whitehead JM, Wilken JM. Step-to-step transition work during level and inclined walking using passive and powered ankle-foot prostheses. *Prosthet Orthot Int*. 2016 Jun;40(3):311-9.
107. Samitier CB, Guirao L, Costea M, Camós JM, Pleguezuelos E. The benefits of using a vacuum-assisted socket system to improve balance and gait in elderly transtibial amputees. *Prosthet Orthot Int*. 2014 Sep 26.
108. Scheck and Siress Advanced Orthotics and Prosthetics. Below knee prosthesis. Vacuum Assisted Socket System (VASS™). Accessed December 2, 2019. Available at URL address: <http://www.scheckandsiress.com/?s=below+knee+prosthesis>
109. Scheme E., Englehart K. (2011). Electromyogram pattern recognition for control of powered upper-limb prostheses: state of the art and challenges for clinical use. *J. Rehabil. Res. Dev.* 48, 643–660.
110. Segal AD, Orendurff MS, Klute GK, McDowell ML, Pecoraro JA, Shofer J, Czerniecki JM. Kinematic and kinetic comparisons of transfemoral amputee gait using C-Leg(R) and Mauch SNS(R) prosthetic knees. *J Rehabil Res Dev*. 2006 Nov-Dec;43(7):857-70.
111. Segal AD, Zelik KE, Klute GK, Morgenroth DC, Hahn ME, Orendurff MS, Adamczyk PG, Collins SH, Kuo AD, Czerniecki JM. The effects of a controlled energy storage and return prototype prosthetic foot on transtibial amputee ambulation. *Hum Mov Sci*. 2011 Nov 17.

112. Selvarajah K, Datta D. An unusual complication of a myoelectric prosthesis. *Prosthet Orthot Int*. 2001 Dec 1;25(3):243-5.
113. Seymour R, Engbretson B, Kott K, Ordway N, Brooks G, Crannell J, Hickernell E, Wheeler K. Comparison between the C-leg microprocessor-controlled prosthetic knee and non-microprocessor control prosthetic knees: a preliminary study of energy expenditure, obstacle course performance, and quality of life survey. *Prosthet Orthot Int*. 2007 Mar;31(1):51-61.
114. Sherman RA, Casey Jones DE. The amputees guide to the amputation and recovery process. Second edition, 1995. Accessed December 2, 2019. Available at URL address: <http://www.fizjoterapeutom.pl/files/13/amputeeguide.pdf>
115. Silcox DH 3rd, Rooks MD, Vogel RR, Fleming LL. Myoelectric prostheses. A long-term follow-up and a study of the use of alternate prostheses. *J Bone Joint Surg Am*. 1993 Dec;75(12):1781-9.
116. Simon AM, Fey NP, Ingraham KA, Finucane SB, Halsne EG, Hargrove LJ. Improved Weight-Bearing Symmetry for Transfemoral Amputees During Standing Up and Sitting Down With a Powered Knee-Ankle Prosthesis. *Arch Phys Med Rehabil*. 2016 Jul;97(7):1100-6.
117. Son H-S, Yildirim T, Khoramnia R, et al. Implantation of a small-aperture intraocular lens and a partial aniridia implant in eyes with traumatic iris defects. *Am J Ophthalmol Case Rep*. 2020 Mar;18:100673.
118. Stein RB, Walley M. Functional comparison of upper extremity amputees using myoelectric and conventional prostheses. *Arch Phys Med Rehabil*. 1983 Jun;64(6):243-8.
119. Struchkov V, Buckley JG. Biomechanics of ramp descent in unilateral trans-tibial amputees: Comparison of a microprocessor controlled foot with conventional ankle-foot mechanisms. *Clin Biomech (Bristol, Avon)*. 2016 Feb;32:164-70.
120. Theeven PJ, Hemmen B, Geers RP, Smeets RJ, Brink PR, Seelen HA. . Influence of advanced prosthetic knee joints on perceived performance and everyday life activity level of low-functional persons with a transfemoral amputation or knee disarticulation. *J Rehabil Med* 2012; 44(5):454-61.
121. Tintle SM, Baechler MF, Nanos GP 3rd, Forsberg JA, Potter BK. Traumatic and trauma-related amputations: Part II: Upper extremity and future directions. *J Bone Joint Surg Am*. 2010 Dec 15;92(18):2934-45.
122. Touch Bionics Inc. i-Limb Ultra Hand. Accessed October 10, 2017. Available at URL address: <http://www.touchbionics.com/products/active-prostheses/i-limb-ultra/>
123. Trallesi M, Delussu AS, Fusco A, Iosa M, Aversa T, Pellegrini R, Brunelli S. Residual limb wounds or ulcers heal in transtibial amputees using an active suction socket system. A randomized controlled study. *Eur J Phys Rehabil Med*. 2012 May 28.
124. Uellendahl JE. Upper extremity myoelectric prosthetics. *Phys Med Rehabil Clin N Am*. 2000 Aug;11(3):639-52.
125. United States Department of Veterans Affairs. Computerized lower limb prosthesis. Veterans Health Administration. Office of Research and Development. Health Service Research and Development Service. Management Decision and Research Center. Technology Assessment Program. Short report no.2. Boston (MA): MDRC; 2000 Mar. Accessed August 6, 2014. Available at URL address: <http://www4.va.gov/VATAP/Publications.asp>
126. United States Department of Veterans Affairs. VA Technology Assessment Program. VHA Prosthetic and Sensory Aids Service Strategic Healthcare Group's Prescribing Guidelines for Computerized Lower Extremity Prosthesis. September 13, 2000. Archived.

127. U.S. Food and Drug Administration (FDA). Center for Device and Radiological Health (CDRH). Code of Federal Regulations. Title 21, Volume 8. Part 890-Physical Medicine Devices. Subpart D-Physical Medicine Prosthetic Devices. Updated April 1, 2016. Accessed December 2, 2019. Available at URL address: <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfcfr/CFRSearch.cfm?fr=890.3420>
128. United States Food and Drug Administration. Center for Devices and Radiological Health. CustomFlex™ Artificial Iris. Premarket Approval P170039. May 30, 2018. Accessed December 4, 2019. Available at URL address: https://www.accessdata.fda.gov/cdrh_docs/pdf17/P170039A.pdf
129. United States Food and Drug Administration. Center for Devices and Radiological Health. Humanitarian Device Exemption (HDE). The Osseoanchored Prostheses for the Rehabilitation of Amputees (OPRA). H080004. July 16, 2015.
130. van der Niet O, Bongers RM, van der Sluis CK. Functionality of i-LIMB and i-LIMB pulse hands: case report. J Rehabil Res Dev. 2013;50(8):1123-8.
131. Washington State Department of Labor and Industries. Office of Medical Director. Health Technology Assessment Brief. Otto Bock Harmony Vacuum Assisted Socket System (VASS). Updated 2003 Apr 3. Accessed July 24, 2017. Available at URL address: <http://www.lni.wa.gov/ClaimsIns/Files/OMD/OttoBockVASS.pdf>
132. Washington State Department of Labor and Industries. Office of Medical Director. Health Technology Assessment. Microprocessor-controlled prosthetic knees. Revised 2002 Aug 16. Accessed August 6, 2014. Available at URL address: <http://www.lni.wa.gov/ClaimsIns/Providers/TreatingPatients/ByCondition/CovMedDev/SpecCovDec/Microprocessor.asp>
133. Weissbart SB, Ayres BD. Management of aniridia and iris defects: An update on iris prosthesis options. Curr Opin Ophthalmol. 2016 May;27(3):244-249.
134. Wolf EJ, Everding VQ, Linberg AA, Czerniecki JM, Gambel CJ. Comparison of the Power Knee and C-Leg during step-up and sit-to-stand tasks. Gait Posture. 2013 Jul;38(3):397-402.
135. Wolf SI, Alimusaj M, Fradet L, Siegel J, Braatz . Pressure characteristics at the stump/socket interface in transtibial amputees using an adaptive prosthetic foot. Clin Biomech (Bristol, Avon). 2009 Dec;24(10):860-5.
136. Work Loss Data Institute. Shoulder (acute & chronic). Corpus Christi (TX): Work Loss Data Institute; 2008. 217 p
137. Worksafe. Workers Compensation Board of BC. Evidence Based Practice Group. A review of microprocessor-controlled knee prostheses. 2003 Nov 27. Accessed July 30, 2015. Available at URL address: <http://www.lni.wa.gov/ClaimsIns/Providers/TreatingPatients/ByCondition/CovMedDev/SpecCovDec/Microprocessor.asp>
138. Yoeruek E, Bartz-Schmidt KU. A new knotless technique for combined transscleral fixation of artificial iris, posterior chamber intraocular lens, and penetrating keratoplasty. Eye (Lond). 2019 Mar;33(3):358-362.
139. Zachariah SG, Saxena R, Ferguson JR, Sanders JE. Shape and volume change in the transtibial residuum over the short term: preliminary investigation of six subjects. J Rehabil Res Dev. 2004 Sep;41(5):683-94.

140. Zeilig G, Weingarden H, Zwecker M, Dudkiewicz I, Bloch A, Esquenazi A. Safety and tolerance of the ReWalk™ exoskeleton suit for ambulation by people with complete spinal cord injury: a pilot study. J Spinal Cord Med. 2012 Mar;35(2):96-101.

“Cigna Companies” refers to operating subsidiaries of Cigna Corporation. All products and services are provided exclusively by or through such operating subsidiaries, including Cigna Health and Life Insurance Company, Connecticut General Life Insurance Company, Cigna Behavioral Health, Inc., Cigna Health Management, Inc., QualCare, Inc., and HMO or service company subsidiaries of Cigna Health Corporation. The Cigna name, logo, and other Cigna marks are owned by Cigna Intellectual Property, Inc. © 2021 Cigna.